

Response To Pre-Bid Queries			
GEM/2022/B/2093051			
OT Table with Orthopedic attachment			
Tender Para	TENDER SPECIFICATION	RESPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
5	Should be able to slide longitudinally more than 250mm both side	We suggest please delete this point, as <b>for Orthopedic procedures, longitudinal slide is of no clinical relevance.</b>	<b>Amended as:- DELETED</b>
7	Two foldable and detachable radiolucent traction bars fixed beneath the seat plate with two adjustable pivot joints.	We suggest Radiolucent traction bars to be detachable only, please delete foldable.	<b>Amended as:-</b> Two detachable radiolucent traction bars fixed beneath the seat plate with two adjustable pivot joints.
12f.	Battery capacity for approximately 2 weeks with average use	We suggest, Battery capacity for 1 week is sufficient for all practical purposes.	Amended as:- Battery capacity for approximately <b>01</b> weeks with average use
12g.	Can be operated directly from the mains for all electro hydraulic and Manual override movements	We suggest, Electrical override in place of Manual Override.	no change considered
13e.	Traction bars radiolucent-02	We suggest to <b>include trolleys for radiolucent traction bars.</b>	<b>Added para:- 01 trolley for radiolucent traction bars.</b>
13h.	Traction boot small pair with multiplanar rotation	We suggest a universal size for traction boot, hence to be deleted.	<b>Amended as:- DELETED</b>