

Response To Pre-Bid Queries			
GEM/2022/B/2092386 Dated: 07.04.2022			
Full Fowler electrical beds with accessories			
Tender Para	TENDER SPECIFICATION	RESPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
	Full Fowler electrical beds with accessories	Please amend as below to avoid ambiguities:"Fully Motorized ICU bed with accessories"	Amended as:- Fully Motorized ICU bed with accessories
4	Profile frame with easy removable ABS plastic covers	Is Five section Poly Propylene detachable top is required on four section top bed frame? Kindly clarify our requirement	Amended as:- 04 section table top with 4-5" thick multilayered mattress.
7	Electrostatic painted metal fraine	Request to please amend this specification as below: foxy powder coated metal frame"	Amended as:- Electrostatic painted metal fraine/ foxy powder coated metal frame
11	Plastie crash bumpers	Request to please amend this specification as below: "Four corner rubber buffers of 100 M dia." Which is more durable	Amended as:- Plastic/rubber crash bumpers
15	Height range: 47 cm	There will be confusion in this specification hence please amend specification as below: Height adjustable: Min. 465 mm To Max.830 mm H JWithout Mattress)	Amended as:-Height Range:- 470 to 770 mm (without mattress) with tolerance of +/- 5%
22	Should be supplied with other essential accessories	Kindly clarify which accessories are required in ICU Bed to offer most suitable and economical product.	1. 4-5" thick multilayered mattress - 1 no 2. Antimicrobial and flame resistant mattress cover - 1 no 3. SS 304 grade saline stand - 2 nos 4. Urinary bag/drain bag holder - 1 no
23	Manufacturer should be ISO 13485 certified		Manufacturer should have ISO 13485 certification issued from: Ay certificatio bodies registered with NABCB under Medical Devices quality maagemet system OR Ay notified body registered with CDSCO OR Any 4 digit European CE notified body