

Response To Pre-Bid Queries			
GEM/2022/B/2098655 Dated: 08-04-2022			
Cavitron Ultrasonic Aspirator (CUSA)			
Tender Para	TENDER SPECIFICATION	RESPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
3.6	It should have on and button	It should have on, off and stand by button. The stand by button protects from any damage if footswitch is pressed by mistake.	No change considered
3.10	Option B (Resuable/disposable accessories) Two(2) Universal handpieces with all mentioned tips including bone sculpting (shape & sizes as mentioned in option A) should be supplied for Dissection Process For reusable tips (made of titanium) undertaking from Manufacturer should be submitted stating that the tips are autoclavable and reusable. Also all the quoting firms should confirm in writing in their offer that they guarantee free of cost supply of Handpiece tips (reusable) for guarantee/ warranty period of five years. If Disposable tips are being offered then Two universal handpieces with 30 tips each mentioned including bone sculpting (shape & sizes as mentioned in option A) (ie. total 150 tips in total). (BOQ shall be revised accordingly)	Option B (Resuable/disposable accessories) Two(2) Universal handpieces of different frequency --one each of high frequency (34-36kHz) and low frequency (23-25 kHz) with all below mentioned tips including bone sculpting should be supplied for Dissection Process . For reusable tips (made of titanium) undertaking from Manufacturer should be submitted stating that the tips are autoclavable and reusable. The quantity of various tips as mentioned below: <ul style="list-style-type: none"> • Tip for Soft Tissue (Precision Curved)- 4 Nos • Tip for Soft Tissue (Standard Long Curved)- 4 Nos. • Tip for Soft Tissue (Standard Curved)- 4Nos. • Micro Tip for Soft Tissue (Curved) – 4 Nos. • Bone Sculpting- 2 No. 	Option B (Resuable/disposable accessories) Two(2) Universal handpieces with all mentioned tips including bone sculpting (shape & sizes as mentioned in option A) should be supplied for Dissection Process For reusable tips (made of titanium) undertaking from Manufacturer should be submitted stating that the tips are autoclavable and reusable. Also all the quoting firms should confirm in writing in their offer that they guarantee free of cost supply of Handpiece tips (reusable) for guarantee/ warranty period of five years. If Disposable tips are being offered then Two universal handpieces with 30 tips each mentioned including bone sculpting (shape & sizes as mentioned in option A) (ie. total 150 tips in total).
	Also all the quoting firms should confirm in writing in their offer that they guarantee free of cost supply of Handpiece tips (reusable) for guarantee/ warranty period of five years.	Please delete " Also all the quoting firms should confirm in writing in their offer that they guarantee free of cost supply of Handpiece tips (reusable) for guarantee/warranty period of five years" Though these are reusable but due to wear and tear ,tips cannot be guaranteed/warranted free of cost for Five years.	No change considered
	suggested by bidder	Should have tissue select/Patient Safety dedicated control for differentiating between firm and tougher tissues and patient safety. Advantage of preserving pia, vital brain arteries , cranial nerves during skull base meningioma, glioma resection, which is very important for skull base tumor operations.	Not considered