

Response To Pre-Bid Queries			
GEM_2022_B_2116323_dated- 26.4.22.			
Platelet Apheresis Machine			
Tender Para	TENDER SPECIFICATION	RESPRESENTATION RECEIVED FROM THE FIRMS	COMMITTEE RECOMMENDATION
	MII Preference clause	As there are currently no manufactures of apheresis device in india, we hereby request the committee to kindly remove this point.	No change considered
12	Extracorporeal volume 150-250ml.	The Extracorporeal volume range should be increased up to – 290ml to accommodate all therapeutic procedures.	No change considered
	suggested by bidder	Blood flow should have range 10ml-120 ml per minute to accommodate pediatric patients. Reason- To accommodate the pediatric patients for procedure with low blood volume.	Added para:- Blood flow should have range 10ml-120 ml per minute
	suggested by bidder	System should have custom prime recommendation to accommodate low TBV patients. Reason- Patient with low (HCT) or pediatric patient may experience problems with higher ECV, to keep the patient safe we can do advance custom priming/blood priming based on recommendation from the device.	No change considered
	suggested by bidder	System should be able to do platelet flush by monitoring buffy coat thickness in order reduce platelet loss during TPE procedures. Reason- This is vital for point for ill patients undergoing TPE, the wastage of platelets shall be avoided by applying platelet flush option	No change considered