

12.05.2021

**Amendment No. 7****Sub: Amendment to the referred tender enquiry****Ref.: Tender Enquiry HITES/PCD/AIIMS-IV/33/Nursing/19-20 Dated 11-03-2020 read with Notice on Extension of Due Dates dated 25.03.2020 read with Amendment no 1 , 2, 3, 4, 5 & 6 dated 03.03.2021, 25.03.2021, 12.04.2021, 23.04.2021, 03.05.2021 & 04.05.2021 respectively.**

The following changes are being incorporated in the above referred Tender Enquiry Document.

**SECTION VII****Technical Specifications**

<b>TECHNICAL AMENDMENT</b>		
<b>NIB Ref: HITES/PCD/AIIMS-IV/33/Nursing/19-20 dated 11.03.2020</b>		
<b>Item No. 10 Full Body Pregnancy Simulator (2021_HLL_72102_10)</b>		
<b>Tender Page &amp; Para</b>	<b>TENDER SPECIFICATION</b>	<b>AMENDED AS</b>
Amndt 3, pg 8, Para II. 9	Should support and detect McRoberts Maneuver	Should support McRoberts Maneuver
Amndt 3, pg 8, Para II. 10	Should support and detect suprapubic pressure with palpable symphysis pubis	Should support suprapubic pressure with palpable symphysis pubis
Amndt 3, pg 8, Para II. 11	Should simulate realistic shoulder dystocia due to realistic pelvic bone movement and it should be resolved once posterior arm is extracted	Should be able to simulate shoulder dystocia
Amndt 3, pg 9, Para III.b	b. Should allow Zavanelli maneuver with detection and logging in event log	Should allow Zavanelli maneuver possibility for logging in event log
Amndt 3, pg 9, Para III.c	c. Trendelenburg position with detection and logging in event log	Trendelenburg position with possibility for logging in event log
Amndt 3, pg 9, Para III.d	d. Left lateral tilt with detection and logging in event log	Left lateral tilt with possibility for logging in event log
Amndt 3, pg 9, Para 13	Should have Fetal heart sounds – 4 locations based on fetal presentation	Should have Fetal heart sounds
Amndt 3, pg 9, Para 14	Should have clinically accurate fetal size with tactile realism – 5th percentile on the WHO growth chart	Should have clinically accurate fetal size with tactile realism

Amndt 3, pg 9, Para 14.d	d. Should provide fetal neck traction graph to give real time feedback on force applied during vacuum or forcep delivery	Deleted
Amndt 3, pg 9, Para 14.f	f. Fetus should have audible cry upon delivery	Should be able to simulate fetal cry
Amndt 3, pg 9, Para 14.g	g. Should display predicted 1-minute and 5-minute APGAR scores based on integrated maternal-fetal physiology	Should be able to display 1-minute and 5-minute APGAR scores
Amndt 3, pg 9, Para 14.j	j. Should have facility bimanual compression and uterine massage detection	Should have facility bimanual compression and uterine massage
Amndt 3, pg 10, Para VI. 10	Should have advance CPR matrices	Should be compatible for CPR
Amndt 3, pg 10, Para b.3	Should have bilateral pulses: carotid, radial, and dorsalis pedis; with controllable pulse strength	Should have palpable pulses, with controllable pulse strength
Amndt 3, pg 10, Para c.4	Should allow Bilateral NIBP measurement	Should allow NIBP measurement
Amndt 3, pg 10, Para d.1	Should simulate seizures with rhythmic movement of arms, rapid blinking, and jaw movement	Should be able to simulate seizures
Amndt 3, pg 10, Para d.2	Should have Reactive pupils	Should be able to simulate pupillary response
Amndt 3, pg 10, Para d.3	Should have facility to change the colour according to condition of the patient like yellow for jaundice, red in case of blood clot	Deleted
Amndt 3, pg 10, Para d.4	Should have Blinking eyes	Deleted
Amndt 3, pg 10, Para e.1	Should have inbuilt postpartum bleeding tank at-least (1,800 ml)	Should have inbuilt postpartum bleeding tank
Amndt 3, pg 10, Para e.2	Should have Bilateral IV arms with realistic flashback	Should have bilateral iv arms
Amndt 3, pg 11, Para C	Pre programmed clinical scenarios for non-gravid patient should include :	Pre programmed clinical scenarios for non-gravid patient should include : (Preferable)

Amndt 3, pg 11, Para C.a	a Chronic Heart failure exacerbation	a Chronic Heart failure exacerbation (Preferable)
Amndt 3, pg 11, Para C.b	b. Acute Respiratory Distress Syndrome	b. Acute Respiratory Distress Syndrome (Preferable)
Amndt 3, pg 11, Para C.c	c Sepsis with Hypotension	c Sepsis with Hypotension (Preferable)
Amndt 3, pg 11, Para C.d	d. Brain Attack with Thrombolytic Therapy	d. Brain Attack with Thrombolytic Therapy (Preferable)
Amndt 3, pg 11, Para C.e	e Motor Vehicle Collision with Hypovolemic Shock	e Motor Vehicle Collision with Hypovolemic Shock (Preferable)

**All other contents of the Tender enquiry including terms & conditions remain unaltered.  
Note:**

- I. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**