

19-03-2022

Amendment no. 2

Sub: Technical Specification Amendment for the tender of Mother & Child Block -AIIMS , New Delhi-110029.

Ref.: GeM Bid No GEM/2022/B/1943446 Dated: 10-02-2022 for item 'Vacuum Assisted Delivery System (Qty: 03 nos)'

The following changes are being incorporated in the above referred GeM Bid Number only.

Existing:

Bid Details	
Bid End Date/Time	23.03.2022,15.00 hrs
Bid Opening Date/Time	23.03.2022,15.30 hrs

Read as:

Bid Details	
Bid End Date/Time	29.03.2022,15.00 hrs
Bid Opening Date/Time	29.03.2022,15.30 hrs

Sr. No.	Page & Para	EXISTING TENDER SPECIFICATION	READ AS
1	Pg 1 Para 5	It should have autoclavable double jars (2.5ltr.) of unbreakable PSU material with disposable collection system with highly burst resistant and double layers liners.	It should have autoclavable double jars (2.5ltr.) of unbreakable PC (Poly Carbonate) / unbreakable PSU material with disposable collection system with highly burst resistant and double layers liners.

2	Pg 1 Para 17	It should be USFDA approved.	Equipment Should be European CE with four digit notified body number or USFDA or BIS approved for the quoted model and certificate to be submitted.
3	Pg 1 Para 18	It should also have certification of ISO 9001, ISO 13485, CE 0123	Manufacturer should have ISO 13485 certification.

All other contents of the Bid Document including terms & conditions remain unaltered.

Note:

Prospective Bidders are also advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid.