

19-03-2022

Amendment no. 2

Sub: Technical Specification Amendment for the tender of Mother & Child Block -AIIMS , New Delhi-110029.

Ref.: GeM Bid No GEM/2022/B/1962669 Dated: 17-02-2022 for item 'LOW END – MID END VENTILATOR (Qty: 02 nos)'

The following changes are being incorporated in the above referred GeM Bid Number only.

| Sr. No. | ATC Page & Para | EXISTING SPECIFICATION | READ AS |
|----------------|----------------------------|--|---|
| 1 | Pg 2 Para 11 c) | Display user defined trends (tabular and graphical) for 48-72 hrs. | Display user defined trends (tabular and graphical) for 24 hours |
| 2 | Pg 2 Para 16 | Should have inbuilt ultrasonic nebulizer for nebulizing drugs without interrupting mechanical ventilation & should have EtCO2 with capnography waveform display. | Should have inbuilt pneumatic nebulizer capable of delivering <or 3 micron particle for nebulizing drugs without interrupting mechanical ventilation & should have EtCO2 with capnography waveform display. |
| 3 | Pg 2 Para 17 | Should have servo controlled USFDA approved humidifier. | Ventilator and compressor should be European CE with four digit notified body number or US FDA or BIS approved for the quoted model and certificate to be submitted. OR Should meet IEC 60601-1, IEC 6060-1-2 & ISO 80601-2-12 standards and should submit valid test report from NABL accredited labs or from the labs in their country of origin (in case of foreign manufacturer) for the quoted model. |
| 4 | Pg 3 Para 26 | Should be capable of upgrading for charting solutions. | DELETED |

All other contents of the Bid Document including terms & conditions remain unaltered.

Note:

Prospective Bidders are also advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid.