

19-03-2022

**Amendment no. 2**

**Sub: Technical Specification Amendment for the tender of Mother & Child Block -AIIMS , New Delhi-110029.**

**Ref.: GeM Bid No GEM/2022/B/1952018 Dated: 14-02-2022 for item 'CRRT Machine (Qty: 01 no)'**

The following changes are being incorporated in the above referred GeM Bid Number only.

**Existing:**

Bid Details	
Bid End Date/Time	23.03.2022,15.00 hrs
Bid Opening Date/Time	23.03.2022,15.30 hrs

**Read as:**

Bid Details	
Bid End Date/Time	29.03.2022,15.00 hrs
Bid Opening Date/Time	29.03.2022,15.30 hrs

Sr. No.	Page & Para	EXISTING SPECIFICATION	READ AS
1	Pg 1 Para 7	There should be provision of separate pump for pre-blood infusion	There should be provision of separate <b>fifth</b> pump for pre-blood infusion
2	Pg 2 Para 21	The system should have provision of changing syringe size	The system should have provision of changing syringe size ( <b>optional</b> )
3	Pg 2 Para 23	The system should have an inbuilt dosage calculator	The system should have an inbuilt dosage calculator ( <b>optional</b> )
4	Pg 2 Para 35	The system should have provision for recording details of therapy for up to 80 -90 hours, and at least 500 events	The system should have provision for recording details of therapy for <b>at least 48 hours</b> , and at least <b>100</b> events

5	Pg 3Para 43	The system should have battery backup of at least 60 minutes	The system should have battery backup of at least <b>15 minutes</b>
---	-------------	--	---

**All other contents of the Bid Document including terms & conditions remain unaltered.**

**Note:**

Prospective Bidders are also advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid.