

19-03-2022

Amendment no. 2

Sub: Technical Specification Amendment for the tender of Mother & Child Block -AIIMS , New Delhi-110029.

Ref.: GeM Bid No GEM/2022/B/1962740 Dated: 17-02-2022 for item 'Warming blanket with hot air warming system under GEM category Patient Warming System (Qty: 29 nos)'

The following changes are being incorporated in the above referred GeM Bid Number only.

Sr. No.	ATC Page & Para	EXISTING SPECIFICATION	READ AS
1	Pg 1 Para 4	Should have short warm up time of less than 40 seconds.	Should have short warm up time within 1 minute
2	Pg 1 Para 5	Should have accuracy of +/- 1 deg C, supported by authenticated paper.	Should have accuracy of +/- 1.5 deg C , supported by authenticated paper.
3	Pg 1 Para 8	Should have automatic step-down facility from high setting to medium setting after 45mins.	Should have automatic step-down facility from high setting to medium setting between 40-60 mins.
4	Pg 1 Para 13	Should have safety alarms like over temperature, power disconnection.	Should have safety alarms like over temperature, hose disconnection
5	Pg 1 Para 19	Should have USFDA/BIS/EU-CE.	Should have USFDA/BIS/EU-CE or should meet IEC 60601-1, IEC 60601-1-2 & IEC 80601-2-35 standard and should submit valid test report from any NABL accredited labs or from the labs in their country of origin (in case of foreign manufactures) for the quoted model

All other contents of the Bid Document including terms & conditions remain unaltered.

Note:

Prospective Bidders are also advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid.