

19-03-2022

Amendment no. 2

Sub: Technical Specification Amendment for the tender of Mother & Child Block -AIIMS , New Delhi-110029.

Ref.: GeM Bid No GEM/2022/B/1955584 Dated: 17-02-2022 for item 'CTG Machine with Central Monitor (Qty: 10 nos)'

The following changes are being incorporated in the above referred GeM Bid Number only.

Sr. No.	ATC Page & Para	EXISTING TENDER SPECIFICATION	READ AS
1	Pg 1 Para 1	Twins Monitoring should be Standard and upgradable to triplets monitoring.	It should have Wireless twins Monitoring as Standard and preferably upgradable to triplets monitoring.
5	Pg 1 Para 5	Should have Minimum 6.5 Inch Color TFT High Resolution Display.	Should have Minimum 12 Inch Color TFT High Resolution Display.
6	Pg 1 Para 6	Should Have Touch Screen Facility	Should have Rotary Knob/ Touch Screen Facility
7	Pg 1 Para 7	Should be upgradable to Maternal ECG and NIBP	It should have Maternal ECG, NIBP & SpO2 as a standard feature.
12	Pg 1 Para 15	Upgradable to Interface to Fetal Telemetry System	Deleted
13	Pg 1 Para 16	Water Proof transducers with LED Indicators	Deleted
14	Pg 1 Para 17	Should have automatic Channel Management so that the fetal monitor can detect any Transducer in any socket.	Deleted
15	Pg 1 Para 19	Wall Mount facility should be available	Wall mount stand should be supplied as standard.
16	Pg 1	Must be a US FDA and CE approved	Equipment Should be European CE with four digit

	Para 21	equipment	notified body number or USFDA or BIS approved for the quoted model and certificate to be submitted.
17	Pg 1 Para 1	Fetal Monitor Central charting and viewing station: It should support a maximum of 250 monitors connectivity	Fetal Monitor Central charting and viewing station: It should support a minimum of 16 monitors connectivity with software license.
18	Pg 1 Para 2	Fetal Monitor Central charting and viewing station: It should support a maximum of 250 monitors connectivity.	Fetal Monitor Central charting and viewing station: It should support a minimum of 16 monitors connectivity.
19	Pg 1 Para 3	Fetal Monitor Central charting and viewing station: It should provide viewing access of data through tablets.	Deleted
20	Pg 1 Para 5	Fetal Monitor Central charting and viewing station: It should record Fetal Heart Rate (up to 3 at a time), Toco, Intrauterine Pressure, Direct Fetal Heart Rate, Maternal ECG, Maternal pulse, SpO2, Temperature and NBP	Fetal Monitor Central charting and viewing station: It should record Fetal Heart Rate (up to 3 at a time), Toco, Direct Fetal Heart Rate, Maternal ECG, Maternal pulse, SpO2, Temperature and NBP
22	Pg 2 Para 8	Fetal Monitor Central charting and viewing station: It should have remote view and control over monitor.	Deleted
23	Pg 2 Para 10	Fetal Monitor Central charting and viewing station: It should have capability to enter notes and vitals on trace.	Deleted
24	Pg 2 Para 12	Fetal Monitor Central charting and viewing station: It should display notes and events marked on monitor.	Deleted
25	NA	NA	Added Para: Details of required hardware & software is added
26	NA	NA	Added Para:

			Water proof Wireless probes for TOCO, FHR & Moveme set Each)
27	NA	NA	<p>Added Para:-</p> <ol style="list-style-type: none"> 1. Server with minimum configuration of (Intel Xenon 2*1.5 GHZ processor 12C Silver, 8 TB storage, 128GM RAM) along with latest compatible Windows Server OS : 1 No 2. Central Station with All in ONE PC with minimum 21" Display (RAM: 16GB, Processor Core i5, Storage: 120 GB) : 1 No 3. LAN cabling between CTG & server to be done by the bidder and price to be quoted separately. 4. Any other hardware/software required for CTG to be supplied without any additional cost

All other contents of the Bid Document including terms & conditions remain unaltered.

Note:

Prospective Bidders are also advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid.