

19-03-2022

Amendment no. 2

Sub: Technical Specification Amendment for the tender of Mother & Child Block -AIIMS , New Delhi-110029.

Ref.: GeM Bid No GEM/2022/B/1943386 Dated: 10-02-2022 for item 'Digital Video Colposcope (Qty: 01 no)'

The following changes are being incorporated in the above referred GeM Bid Number only.

Existing:

Bid Details	
Bid End Date/Time	23.03.2022, 15.00 hrs
Bid Opening Date/Time	23.03.2022, 15.30 hrs

Read as:

Bid Details	
Bid End Date/Time	29.03.2022, 15.00 hrs
Bid Opening Date/Time	29.03.2022, 15.30 hrs

Sr. No.	Page & Para	EXISTING TENDER SPECIFICATION	READ AS
1	Pg 1 Para 1	Should have > 1000 TV Lines in Std. mode and > 1200 TV Lines in Gamma mode.	Deleted
2	Pg 1 Para 4	Should have aspect ratio of 16:9 (HD) for cinema like picture	Should have aspect ratio of 16:9 (Full HD)
3	Pg 1 Para 5	The video colposcope must have magnification from min. 1x to max. 55x	The video colposcope must have magnification from min. 1-50x or above.
4	Pg 1	Facility to increase and decrease the	Deleted

	Para 6	light intensity	
5	Pg 1 Para 7	Varying image color (5 different settings)	Deleted
6	Pg 1 Para 8	Should have 20,00,000 No. of pixels	Deleted
7	Pg 1 Para 9	Should have Full HD component output: Y Pb Pr., DVI	Should have Full HD component output: Y Pb Pr./ DVI/HDMI/ USB 3.0
8	Pg 1 Para 10	Should have analog output: BNC and SVHS	Deleted
9	Pg 1 Para 12	High MCD super bright white shadow less LED light for true color reproduction.	Deleted
10	Pg 1 Para 18	There must be 1 step blue filter in the hand-held unit without decrease in illumination.	Deleted
11	Pg 1 Para 22	It should be equipped with Gamma Processor to enhance vascular structure	Deleted
12	Pg 1 Company should provide Colposcopy Image Management software with computer with following facilities (Qty:-1) Para 3	Should have facility to generate hysteroscopy and Cryo-Surgery report	Should have facility to generate Colposcopy and Cryo-Surgery report

13	Pg 2 Company has to provide Desktop Computer with: (One Set) Para 4	LCD Monitor of 20"	LCD Medical Grade Monitor of atleast 24" .
14	Pg 2 Company has to provide Desktop Computer with: (One Set) Para 5	Laser Printer	It should have Color Laser printer.
15	Pg 2 Company has to provide Desktop Computer with: (One Set) Para 6	Computer table	Computer table with mobile trolley with lockable wheels
16	Pg 2	Company should provide: minimum 24" Branded TV Monitor	Company should provide: minimum 24" Medical Grade color Monitor
17	Pg 2 Company should provide: minimum 24" Branded TV Monitor Para 1	Should have resolution of 1920 x 1080	Should have Full HD medical grade monitor

All other contents of the Bid Document including terms & conditions remain unaltered.

Note:

Prospective Bidders are also advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid.