

19-03-2022

Amendment no. 2

Sub: Technical Specification Amendment for the tender of Mother & Child Block -AIIMS , New Delhi-110029.

Ref.: GeM Bid No GEM/2022/B/1933858 Dated: 14-02-2022 for item 'Plasmapheresis Machine (Portable) (Qty: 01 no)'

The following changes are being incorporated in the above referred GeM Bid Number only.

Existing:

Bid Details	
Bid End Date/Time	23.03.2022, 15.00 hrs
Bid Opening Date/Time	23.03.2022, 15.30 hrs

Read as:

Bid Details	
Bid End Date/Time	29.03.2022, 15.00 hrs
Bid Opening Date/Time	29.03.2022, 15.30 hrs

Sr. No.	Page & Para	EXISTING SPECIFICATION	READ AS
1	Pg 1 Para 14	The system should provide blood prime recommendations for the safety of pediatric and anemic patients	The system should have provision for blood prime for the safety of pediatric and anemic patients
2	Pg 1 Para 15	The system should have provision to initiate boluses and record the volumes.	The system should have provision to provide boluses or infusion through return line
3	Pg 1 Para 16	The system should provide a continuous graphical representation of details of ongoing fluid balance	The system should provide a continuous graphical or numerical representation of details of ongoing fluid balance

4	Pg 1 Para 18	The system should have an in-built tube sealer.	The system should have either an inbuilt tube sealer or the option to add on (optional)
5	Pg 1 Para 23	The system should be able to regulate pump speeds based on type of replacement solution.	The system should be able to regulate pump speeds to alter the infusion rate of solution
6	Pg 2 Para 25	The system should be able to operate in ambient temperature of 20-30°C and relative humidity 0-70%.	The system should be able to operate in ambient temperature of 20-25°C and relative humidity 0-70%
7	Pg 2 Para 35	Rates of all consumables should be provided and fixed for a period of 10 years.	List of all essential consumables for use with machine to be provided as annexure (attached), and rates of all these consumables should be provided and fixed for a period of 10 years.

List of Consumables items		
Price of essential consumables as shown below to be quoted separately which shall be considered for bid ranking (the unit price should be fixed for 10 years). These consumables may be ordered as and when required by the user department.		
Sr. No	Item Description	Quantity (approx. qty over 10 years being factored for bid ranking purposes only)
1	Exchange set	1000
2	ACD solution 500 ml	1500

All other contents of the Bid Document including terms & conditions remain unaltered.

Note:

Prospective Bidders are also advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid