

11.05.2021

Amendment No. 9**Sub: Amendment to the referred tender enquiry**

Ref.: Tender Enquiry HITES/PCD/AIIMS-IV/32/Pulmo/19-20 Dated 11-03-2020 read with Notice on Extension of Due Dates dated 25.03.2020 read with Amendment no 1, 2, 3, 4 ,5, 6, 7 & 8 dated 03.03.2021,25.03.2021, 09.04.2021, 12.04.2021, 22.04.2021, 30.04.2021 & 10.05.2021 respectively.

The following changes are being incorporated in the above referred Tender Enquiry Document.

SECTION VII**Technical Specifications**

TECHNICAL AMENDMENT		
NIB Ref: HITES/PCD/AIIMS-IV/32/Pulmo/19-20 dated 11.03.2020		
Item No. 4 - High Resolution Video Bronchoscopy system (2021_HLL_72086_4)		
Tender Page & Para	TENDER SPECIFICATION	AMENDED AS
Pg 52 Para A.7	Compatible with leakage testing device with its air flow and pressure regulation through light source's air pump	Compatible with leakage testing device with its air flow and pressure regulation through light source's air pump or compatible with external leakage testing device

All other contents of the Tender enquiry including terms & conditions remain unaltered.
Note:

- I. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**