

Amendment No. 01

Date- 02-03-2024

Sub: Amendment No.01 to the Tender Enquiry Document

Ref: (i) Tender No: HITES/PCD/SFDJ/TED/23-24/03 dated 17-02-2024

For Tender ID: (2024_HITE_186915_1) : High Dose Rate Brachytherapy System**Section VII**
Technical Specifications

GTE No.: HITES/PCD/SFDJ/TED/23-24/03 (Tender ID: 2024_HITE_186915_1)			
Item Name: High Dose Rate Brachytherapy System			
Sr. No.	Tender document para and pg	Existing Tender Specification	Read As
1	Para 5, Point 5.6 (Applicators for HDR Unit), pg 47	Esophagus applicator – 3 sets (6mm 18Fr, 8mm 24Fr and 12mm 30Fr)	Esophagus applicator – 3 sets (6mm 18Fr/ 10 mm, 8mm 24Fr and 12mm 30Fr).
2	Para 5, Point 5.9 (Applicators for HDR Unit), pg 47	Surface mould – 5 sets for IOHDR applications (Freiberg applicators)	Surface mould – 5 sets for IOHDR applications.
3	Para 5, Point 5.12 (Applicators for HDR Unit), pg 47	The vendor has to supply total 3 sets of transfer tubes (one with main config and 2 sets extra) for both standard fletcher and also for CTMR fletcher. Added Para : The vendor should provide CTMR compatible applicators for cancer cervix : 2 sets.	The vendor has to supply total 3 sets of transfer tubes (one with main config and 2 sets extra) for both standard fletcher and also for CTMR fletcher.
4	NA	Suggestion by bidder	Available AERB approved HDR Room Plan attached
5	NA	Suggestion of HITES	Added Para: Quality Certifications: 1. Equipment should be BIS/USFDA/ European CE approved. 2. Should have import/manufacturing license from Central licensing Authority or State licensing authority of CDSCO for Medical Devices and copy of valid license should be submitted for the quoted model.

Note: If EMD is submitted in the form of BG, then the validity of the BG should be at least 225 days from the date of tender opening.**All other terms and conditions of the tender enquiry remain unaltered**