

Corrigendum 02 in GeM portal

1. Earnest Money Deposit: The Bidders shall submit their EMDs' in proportion to their offered quantity. For instance, if a bidder has to offer 10% of tendered quantity, then the EMD to be submitted shall be 10% of the EMD as per bidding document. The hardcopy instruments of Earnest Money Deposit have to reach the office of HITES, Noida within 5 days of bid submission end date.
2. Supplier shall ensure that the Invoice is raised in the name of Consignee with GSTIN of Consignee only
3. Technical Specification: Performance Parameters: Software: Suitable software for Data Acquisition, analysis of run and also for Gene Expression analysis by relative quantity or normalised expression. Software should have the capacity to analyse data of minimum 10 different runs at a time.

<u>Prebid-Clarifications</u>				
Bid No:- Bid Number: GEM/2020/B/773250 Dated: 03-09-2020				
Item Name:- Real Time PCR (100 Units)			Date of Pre-bid Meeting - 07-09-2020	
Sl. No.	Tender Specification (Clause No & Page No)	Name of the Firm	Representations Received from the bidders	Committee recommendations
1		Bio-Rad Labotratories (India) Pvt. Ltd	kindly clarify, if there are any ceiling for the price bidding in context to the EMD % & EMD amount mentioned in tender document.	No
2			Does the estimated bid value include GST?	Yes
3			Kindly clarify Whether the tender winner can exercise delivery of goods through its authorised agent ?	order shall be placed on the Bidder
4	Performance parameter : Compatible online UPS Data Management system		Whether UPS is required or not and if yes, how many qty. is required 1 or 2.	1 compatible online UPS required with backup of ≥1 hour
5	Performance parameter : Compatible online UPS Data Management system		Do both the UPS are required with 1 hr backup ?	1 compatible online UPS required with backup of ≥1 hour
6	Corrigendum, Page no. 3 point 22		Desktop asked in tender, is it mandatory? Or can a Laptop be supplied in lieu of Desktop?	Technical specification shall prevail
7	page no. 5, under "consignee reporting officer"		Delivery Terms 30 days and location details are requested for planning purpose.	Consignee details for delivery shall be communicated later.
8	STC 1.vi, page no. 5		Bio-Rad doesn't provide services for civil work mentioned in the tender	Technical specification shall prevail
9	STC 1.vi, page no. 5		We will only perform maintenance and repairs for instrument supplied by Bio-Rad only.	Technical specification shall prevail
10	Main tender document, page no.1, Corrigendum Point 11, page 2,		1 Please clarify site addresses ? 2 Inspection before dispatch at our site or post dispatch at consignee site? 3 If dispatch to be done at HITES site @ Noida, who will further dispatch the consignments to end users (institutes) ? 4 In case HITES will dispatch to end user's and any transit damage happens, who will bear the damage cost ?	Consignee details for delivery shall be communicated later.

<u>Prebid-Clarifications</u>				
Bid No:- Bid Number: GEM/2020/B/773250 Dated: 03-09-2020				
Item Name:- Real Time PCR (100 Units)			Date of Pre-bid Meeting - 07-09-2020	
Sl. No.	Tender Specification (Clause No & Page No)	Name of the Firm	Representations Received from the bidders	Committee recommendations
11	Certification of Product CE-IVD	Agilent Technologies India Pvt Ltd	Any Regulatory Body like European CE/IVD/FDA Remarks:- Different vendor have different regulatory bodies certification .So kindly keep option for all in order that reputed Manufacturer like Agilent can also participate.	Already clarified during prebid that the same is covered in the tender specifications.
12	Compatible Sample volume range in micro litre 5- 50		Agilent Reaction Volume range is from 10 to 30 uL. Remarks:_ Since RTPCR is a very sensitive technique and hence, it is always recommended to use lesser volume of sample (To save sample) instead of using higher volumes. Agilent machine can also practically use higher volumes but factory recommendations are always to use less volumes and avoid wasting extra reagent volumes. Manufacturer like Bio Rad & Roche also recommend to use volume range within 10 to 25 uL Only.	The asked volume already falls under the specified range of 5-50 µl.
13	Software: Dedicated licensed full version software for primer and probe design with comprehensive assay design and development guidelines for quantitative and qualitative real time assays to enable designing of custom oligo assays		Please remove this point as software for analysis is already mentioned in earlier tender specs. Primer & probes sequences are readily available through CDC/WHO/ICMR/NIV . Hence there is no requirement for design of primer of probes software specifically for COVID. Since end users intend to use commercially available kits (Covid/Infectious disease) hence there is no requirement for designing new primer and probes as it will be readily available within the kits. Remarks:_ Primer & Probe designed software is intended to use mostly for research applications. Further to software the designed primer and probes require to undergo validation if they require to run in a clinical setup.	Software* (Suitable software for Data Acquisition, analysis of run and also for Gene Expression analysis by relative quantity or normalised expression. Software should have the capacity to analyze data of minimum 10 different runs at a time.
14	Compatible Sample volume range in micro litre. :- 5-50	Thermo Fisher Scientific	Compatible Sample volume range in micro litre. :- 10-50 or more	Sample volume range in micro litre. :- 5-50
15			Request for additional specifications :- Quoted system should have provision for setting up different annealing temperature in same run.:- 5 or more	No Change
16			Request for additional specifications. Quoted system should have variable temperature range.:- 4-99 C or more	No Change