

Amendment no. 2

Sub: Technical Specification Amendment for the tender of Mother & Child Block -AIIMS , New Delhi-110029.

Ref.: GeM Bid No GEM/2022/B/2854526 Dated: 14-12-2022 for item 'Biological Safety Cabinet (Qty: 01 no.)'

The following changes are being incorporated in the above referred GeM Bid Number only.

Sr. No.	Buyer uploaded ATC document	EXISTING TENDER SPECIFICATION	READ AS
1	Pg 1 Para 2	NSF International standard 49 / EN 12469 certified and tested	Deleted
2	Pg 2 Para 23	NA	Added Para: Equipment should have CE / European CE / BIS certification.
3	NA	NA	Added Para: BOQ: 1. Biological Safety Cabinet as specified - 01 No.

Important Note:

1. Prospective Bidders are advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid.
2. The Bid Security/EMD shall be valid for 110 days from Techno – Commercial Tender opening date. The techno-commercial tender opening date shall be considered as mentioned in bid document read with it's amendment(s), if any.
3. In reference to the extension of bid opening date as mention on GeM portal, Participating bidders, who are submitting their Bid security/EMD in the form of Bank Guarantee (BG) or FDR are instructed to extend the validity of their BG/FDR accordingly, i.e. required validity of EMD is to be kept in line with the extended bid opening date.
4. Traders/resellers/distributors/authorized agents will not be considered for availing benefits under PP Policy 2012 for MSEs as per MSE guidelines issued by MoMSME
5. Bidders shall ensure that their tender(s), complete in all respects, are submitted online through **GeM portal only**.