

Amendment No. 07

Sub: Amendment to the referred tender enquiry

Ref.: Tender Enquiry No. HITES/PCD/PMSSY-IV/16/GNS/2021-22 dated 20-07-2021 and its Amendment No. 01 dated 16-Aug-2021, Amendment No. 02 dated 24-Aug-2021 & Amendment No. 03 dated 25-Aug-2021, Amendment No. 04 dated 06-Sep-2021, Amendment No. 05 dated 14-Sep-2021 & Amendment No. 06 dated 21-Sep-2021.

Following changes are being incorporated in the above referred Tender Enquiry Document.

SECTION VII
TECHNICAL SPECIFICATION

For: Tender IDs 2021_HLL_81944_1, 2021_HLL_81944_8 & 2021_HLL_81944_10

Tender Page & Para	EXISTING SPECIFICATION	READ AS
Item No. 1 Multipurpose OT Table		
Para 4.j	Cranial and caudal traversing or longitudinal movement of minimum 200 mm either side	Cranial and caudal traversing or longitudinal movement of minimum 200 mm on both sides.
Para 5.j	Power lift stirrup set with side rail clamp – 1 set Powered lift stirrup set should be remote operated & manually operated	Pneumatic lift stirrup set with side rail clamp - 1 set
Item No. 8 Open Surgery Instruments (Set) for General Surgery		
Para 2.1	All the instruments should be made of corrosion free surgical grade stainless steel SS 400 grade and bidder should submit original OEM certificate for the same	All the instruments should be corrosion free surgical grade SS 400 series stainless steel and bidder should submit original OEM certificate for the same
Item No. 10 Laparoscope Set		
3.8	IMAGE MANAGEMENT SYSTEM	
Para 3.8.f	Latest processor & HDD (atleast 1 TB or more)	Latest processor & HDD (atleast 1 TB or more) Note: <i>It should be an inbuilt system, separate laptop or desktop will not be considered</i>
Para 3.8.n	It should have latest windows operating system & colour laser printer	It should have latest compatible operating system & colour laser printer
	Tissue Grasping forceps – toothed 2x3 teeth Double action jaws, rotating with connector pin for unipolar coagulation, size 5mm, length 33-36cm, dismantling facility, plastic handles with ratchet, autoclavable - 1 no	Tissue Grasping forceps – toothed minimum 2x3 teeth Double action jaws, rotating with connector pin for unipolar coagulation, size 5mm, length 33-36cm, dismantling facility, plastic handles with ratchet, autoclavable - 1 no
	Grasping forceps- Atraumatic Double action jaws, spoon shaped with multiple teeth of jaw and rotating with connector pin for unipolar coagulation, size 5mm, length 33-36cm, dismantling facility, plastic handles without ratchet, autoclavable- 1 no	Grasping forceps- Atraumatic Double action jaws, spoon shaped with multiple teeth of jaw/hollow jaw and rotating with connector pin for unipolar coagulation, size 5mm, length 33-36cm, dismantling facility, plastic handles without ratchet, autoclavable- 1 no
	Fibreoptic Light cables - Fibreoptic Light cables With straight connectors of 4.8mm diameter and 230-250 cm long - 1	Fibreoptic Light cables - Fibreoptic Light cables with straight connectors of 4.2mm or more diameter and 230-250 cm long - 1

Tender Page & Para	EXISTING SPECIFICATION	READ AS
	Fibreoptic Light cables - With straight connectors of 4.8mm diameter and 300 cm long	Fibreoptic Light cables - With straight connectors of 4.2mm or more diameter and 300 cm long
Para 4.1	<p>System as specified. Telescope, Light source, Camera Head, Image Management system and CO2 insufflator should be of same make.</p> <p>All hand instruments should be from same principal manufacturer; incase any of the instrument is not manufactured by the principal manufacturer, the same shall be offered from a second manufacturer; not exceeding 30% of total quantity of hand instruments. However, the bidder is allowed to quote instrument(s) from maximum two manufacturers including the principal manufacturer. (In this case the principal manufacturer should submit an undertaking for non production of the such instrument(s)). Instrument from second manufacturer also should have similar certification as asked in tender and the same should be submitted along with bid</p>	<p>Telescope, Light source, Camera Head, Image Management system and CO2 insufflator should be of same make.</p> <p>Remaining items video cart, suction & irrigation unit, Video Monitor and hand instruments can be from other manufacturers and the manufacturer should have ISO 13485 certification</p>
Para 7.1	Should be US FDA or European CE with 4 digit notified body number or BIS approved product	<p>Should be US FDA or European CE with 4 digit notified body number or BIS approved product (Telescope, Light source, Camera Head, Image Management system and CO2 insufflator)</p> <p>Remaining items video cart, suction & irrigation unit, Video Monitor and hand instruments can be from other manufacturers and the manufacturer should have ISO 13485 certification</p>

Note:

- i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.
- ii. All other contents of the tender enquiry including terms & conditions remain unaltered.
- iii. For Tender IDs 2021_HLL_81944_1, 2021_HLL_81944_8 & 2021_HLL_81944_10: TC compliance sheet v1.1 is being uploaded on CPP Portal for necessary actions of interested Bidders.