

05.01.2022

Amendment No. 4**Sub: Technical amendment to the referred tender enquiry****Ref.: HITES/PCD/AIIMS-IV/46/CTVS/21-22 dated 26.10.2021 read with its Amendment No 1, 2 and 3 dated 26.11.2021, 07.12.2021, 20.12.2021**

The following changes are being incorporated in the above referred Tender Enquiry Document for following item only.

Sl.No	Tender Id	Name of the Item
1	2021_HLL_92629_8	ACT Machine

SECTION I**NOTICE INVITING TENDER (NIT)****2) Tender timeline:****Existing:**

Sl. No.	Description	Schedule
b.	Closing date & time for submission of online bids	05-01-2022, 1400 hrs
c.	Closing date & time for submission of tender processing fee in physical form*	06-01-2022, 1400 hrs
d.	Time and date of opening of online bids	06-01-2022, 1430 hrs

Read as:

Sl. No.	Description	Schedule
b.	Closing date & time for submission of online bids	18-01-2022, 1400 hrs
c.	Closing date & time for submission of tender processing fee in physical form*	19-01-2022, 1400 hrs
d.	Time and date of opening of online bids	19-01-2022, 1430 hrs

SECTION VII
TECHNICAL SPECIFICATIONS

Tender Ref No : HITES/PCD/AIIMS-IV/46/CTVS/21-22 Dated: 26.10.2021

Item Name: ACT Machine (Tender ID : 2021_HLL_92629_8)

Sl. No	Technical Specification	Amended as
3.1	ACT machine having at least two test well	ACT Machine having one test well or more
3.2	The equipment should have 2 channel for measuring the individual clotting time and average difference for the two channels with the help of 2 channel test cartridge	The equipment should have 1 channel or more for measuring the individual clotting time
3.3	Parameters- ACT (Mandatory) APTT & PT (Optional).	Parameters-ACT
3.6	The display should show the coagulation time of 2 channels	The display should show the coagulation time of 1 channel or more

All other contents of the Tender enquiry including terms & conditions remain unaltered.
Note:

- I. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**