

17.01.2023

**Clarification no 2****Sub: Technical Amendment to the referred tender enquiry****Ref.: HITES/PCD/AIIMS-IV/57/Rad-Onco/ 22-23 dated 27-09-2022 read with amendment no 1, 2, 3, 4 and 5 dated 25.10.2022, 05.11.2022, 14.11.2022, 01.12.2022 and 20.12.2020 respectively.**

The following clarification are being incorporated in tender ID 2022\_HLL\_121185\_1 only in the above referred Tender Enquiry Document.

Advanced High Energy Linear Accelerator (Tender ID : 2022_HLL_130211_1)			
Existing Technical Specification			Clarification against the technical para
	Relative Dosimetry Equipment		
2	<b>Radiochromic films :</b> i. Self-developing radiochromic film with sensitivity up to 20 Gy for radiotherapy dosimetry and QA applications shall be provided initial commissioning time and also as and when the department demands for the same. ii. EBT3 Gafchromic film size: 14 x 17 inch EBT3 Gafchromic film size: 8 x 10 inch	250 sheets each	<b>Note: 250 Nos of Radiochromic films, as specified in the technical para will be considered for price ranking purpose. However, the unit price for both the following EBT3 Gafchromic film sizes should be quoted separately.</b> <b>a. 14 x 17 inch</b> <b>b. 8 x 10 inch</b>

**All other contents of the Tender enquiry including terms & conditions remain unaltered.****Note:**

- I. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**