Response To Pre-Bid Queries (Pre-Bid date: 17.11.2022)				
GeM Bid Number: GEM/2022/B/2702566 Dated: 03-11-2022				
Item Name: Mobile X-ray Flat panel DR				
Sr. No.	**Buyer Specification document Para and Page**	**TENDER SPECIFICATION**	**REPRESENTATION RECEIVED FROM THE FIRM**	**COMMITTEE RECOMMENDATION**
1	Para and pg 1	State-of-the-art, battery driven, compact, easily transportable, USFDA approved digital mobile radiographic units with telescopic or articulated arm and flat panel detector are required. It should be suitable for bedside radiography in wards, intensive care units and operation theatres.	To ensure top quality product kindly consider adding : It is mandatory that Radiation causing/controlling component i.e X Ray Generator and X Ray tube ,collimator , software must be from same manufacturer. Attach undertaking. We request you to amend it as State-of-the-art, battery driven, compact, easily transportable, USFDA / European CE approved digital mobile radiographic units with telescopic or articulated arm and flat panel detector are required. It should be suitable for bedside radiography in wards, intensive care units and operation theatres.	State-of-the-art, battery driven, compact, easily transportable, BIS/USFDA/ European CE approved digital mobile radiographic units with telescopic or articulated arm and flat panel detector are required. It should be suitable for bedside radiography in wards, intensive care units and operation theatres.
2	NA	Suggested by bidder	Kindly ADD that Third party US FDA certificate OR US FDA in name of similar/other models will not be accepted.	Not considered
3	NA	Suggested by bidder	BUILT IN CHARGER FOR DETECTORS (To minimize possibility of damage to detector in taking detector to external charging docks)	Not considered
4	NA	Suggested by bidder	LIVE STREAM CAMERA FOR BETTER PATIENT POSITIONING (So operator can observe patient closely and avoid retake due to patient movement)	Not considered
5	Para 6(Other Features), Point g, pg 1	The machine including batteries should not weigh more than 600 kg. Exact weight of the machine to be mentioned.	600 KG unit will be very bulky and such heavy units will be difficult to bring back to department when battery gets discharged. So kindly modify 600 Kg as 430 Kg or less.	Not considered
6	Para 6(Other Features), Point f, pg 1	Entire unit should be USA FDA approved.	Entire unit should be USA FDA / European CE approved	Deleted

Response To Pre-Bid Queries (Pre-Bid date: 17.11.2022)				
GeM Bid Number: GEM/2022/B/2702566 Dated: 03-11-2022				
Item Name: Mobile X-ray Flat panel DR				
Sr. No.	**Buyer Specification document Para and Page**	**TENDER SPECIFICATION**	**REPRESENTATION RECEIVED FROM THE FIRM**	**COMMITTEE RECOMMENDATION**
7	Para 6(Other Features), Point d, pg 1	A grid of minimum 5:1 ratio with size at least 13”xl3” should be supplied.	A grid of a minimum6:1 ratio with suitable size should be supplied	No change considered
8	NA	Suggested by bidder	ONE ADDITIONAL WORKSTATION FOR REPORTING.	Not considered
9	NA	Suggested by bidder	15 Light weight aprons should be provided.	Added Para: 3 Light weight aprons should be provided.
10	NA	Suggested by bidder	15 Thyroid shields to be provided.	Added Para: 3 Thyroid shields to be provided.
11	NA	Suggested by bidder	15 Gonad shield to be provided.	Added Para: 3 Gonad shield to be provided.
12	NA	Suggested by bidder	15 google/eye shield	Added Para: 3 google/eye shield
13	Para , pg 2	State-of-the-art, battery driven, compact, easily transportable, USFDA approved digital mobile radiographic units with telescopic or articulated arm and flat panel detector are required. It should be suitable for bedside radiography in wards, intensive care units and operation theatres	as per prebid discussion	Deleted
14	Para 9, pt c, pg 2	Warranty should be followed by comprehensive annual maintenance contract (CMC) for next five years, covering all items as that of warranty. The Cost of CMC with year wise breakup from 6th to 10th years should be mentioned in the price bid.	as per prebid discussion	Deleted
15	Para 9, pt d, pg 2	95% uptime guarantee should be given. In case downtime exceeds 5%, penalty in the form of extended warranty, double the number of days for which the equipment goes out of service will be applied.	as per prebid discussion	Deleted

Response To Pre-Bid Queries (Pre-Bid date: 17.11.2022)						
GeM Bid Number: GEM/2022/B/2702566 Dated: 03-11-2022						
Item Name: Mobile X-ray Flat panel DR						
Sr. No.	**Buyer Specification document Para and Page**	**TENDER SPECIFICATION**	**REPRESENTATION RECEIVED FROM THE FIRM**	**COMMITTEE RECOMMENDATION**		
16	Para 9, pt e, pg 2	There should be at least three installations of quoted unit in globally. Satisfactory performance report from the existing users should be provided	as per prebid discussion	Deleted		
17	Para 6, pt f, pg 1	Entire unit should be USA FDA approved	as per prebid discussion	Deleted		
18	Para 12, pg 2	NA	as per prebid discussion	**Added Para: Approvals** Should have import/manufacturing license from Central licensing Authority or State licensing authority of CDSCO for Medical Devices and copy of valid license should be submitted for the quoted model.In case the vendor has not yet obtained import/manufacturing license from CDSCO for the quoted model, proof of application for CDSCO medical device license to be submitted in the bid document and valid CDSCO license to be produced at the time of supply/ NOA for the quoted model		
19	Para 12, pg 2	NA	as per prebid discussion	The system should be AERB type approved and the copy of E-LORA Listing should be submitted along with bid. If the quoted model has not been yet installed in India, vendor should submit NOC from AERB.Regular QA according to AERB norms will be responsibility of bidder during warranty and CMC period.		
20	Para 12, pg 2	NA	as per prebid discussion	Regular QA according to AERB norms will be responsibility of bidder during warranty and CMC period.		
21	**Added Para**	NA	as per prebid discussion	**BOQ**		
			1	Mobile X-ray Flat panel DR ,as per the Tender Specification	1	No
			2	DICOM printer	1	No
			3	Light weight aprons	3	Nos

Response To Pre-Bid Queries (Pre-Bid date: 17.11.2022)						
GeM Bid Number: GEM/2022/B/2702566 Dated: 03-11-2022						
Item Name: Mobile X-ray Flat panel DR						
Sr. No.	Buyer Specification document Para and Page	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION		
			4	Thyroid shields	3	Nos
			5	Gonad shield	3	Nos
			6	google/eye shield	3	Nos