Response To Pre-Bid Queries (Pre-Bid date: 17.11.2022)				
GeM Bid Number: GEM/2022/B/2702810 Dated: 03-11-2022				
Item Name:Digital Flat Panel Fluoro cum Radiography				
Sr. No.	**Buyer Specification document Para and Page**	**TENDER SPECIFICATION**	**REPRESENTATION RECEIVED FROM THE FIRM**	**COMMITTEE RECOMMENDATION**
1	Para G (Control console), point 4, pg 1	Added Para : The control software must have latest security windows 10 platform only.	Kindly modify as The control software must have latest security windows 10/equivalent compatible platform.	The control software must have latest security windows 10/equivalent compatible platform.
2	NA	suggested by bidder	Kindly ADD that quoted model must be US FDA approved. Third party US FDA certificate OR US FDA in name of similar/other models will not be accepted.	Not considered
3	Para J (Essential certification), point 3, pg 2	Should have import/manufacturing license from Central licensing Authority or State licensing authority of CDSCO for Medical Devices and copy of valid license should be submitted for the quoted model.In case the vendor has not yet obtained import/manufacturing license from CDSCO for the quoted model, proof of application for CDSCO medical device license to be submitted in the bid document and valid CDSCO license to be produced at the time of supply/ NOA for the quoted model	Pls DELET complete para requirement of CDSCO as it is responsibility of vendor to do customs clearance	No change considered
4	Para I (Accessories), pt 5, pg 2	Radiation protection flaps	Asked flaps are applicable for IITV based unit where IITV mounted on table. Since asked unit is flat panel based in which detector is under the table so protection lead barrier with lead glass is used for radiation protection. Kindly replace radiation protection flaps with protection lead barrier having lead glass.	lead barrier with lead glass for Radiation protection.
5	Para (Site Modification), point 3, pg 2	The cost of site modification for the area of 1000sq.ft and Air-conditioning of Tonnage 20 TR (including standby unit/s) will be considered for Ranking / Evaluation purpose. However, the payment will be paid on pro-rata based on the actual site modification area	as per prebid discussion	The cost of site modification for the area of 350 sq.ft and Air-conditioning of Tonnage 6TR (including standby unit/s) will be considered for Ranking / Evaluation purpose. However, the payment will be paid on pro-rata based on the actual site modification area
6	Paar BOQ(site modification work Work), PG 3, POINT 1	Civil works 1000 Sq.ft	as per prebid discussion	Civil works 350 Sq.ft
7	Paar BOQ(site modification work Work), PG 3, POINT 2	Electrical work 1000 Sq. ft	as per prebid discussion	Electrical work 350 Sq. ft

Response To Pre-Bid Queries (Pre-Bid date: 17.11.2022)				
GeM Bid Number: GEM/2022/B/2702810 Dated: 03-11-2022				
Item Name:Digital Flat Panel Fluoro cum Radiography				
Sr. No.	**Buyer Specification document Para and Page**	**TENDER SPECIFICATION**	**REPRESENTATION RECEIVED FROM THE FIRM**	**COMMITTEE RECOMMENDATION**
8	Paar BOQ(site modification work Work), PG 3, POINT 3	Public health (plumbing and sanitary fittings). 1000 Sq.ft.	as per prebid discussion	Public health (plumbing and sanitary fittings): 350 Sq.ft.
9	Paar BOQ(site modification work Work), PG 3, POINT 4	Air Conditioning 12 TR	as per prebid discussion	Air Conditioning: 06 TR
10	Paar BOQ(site modification work Work), PG 3, POINT 5	Interior Furnishing & Furniture 1000 Sq.ft.	as per prebid discussion	Interior Furnishing & Furniture 350 Sq.ft.
11	Paar BOQ(Miscellaneous), PG 3, POINT 5	Toilet (Optional) - 1	as per prebid discussion	Deleted
12	Para BOQ(Miscellaneous), PG 3	All the Optional items offered should be quoted separately and the price should be valid for warranty period	as per prebid discussion	Deleted
13	NA	NA	as per prebid discussion	Added Para: Indicative Layout drawing attached.