

**Amendment No. 09****Sub: Amendment to the referred tender enquiry**

**Ref.:** Tender Enquiry Document HITES/PCD/PMSSY-IV/25/Mix/2022-23 dated 12-08-2022 and its Amendment No. 01 dated 18-Aug-2022, Amendment No. 02 dated 19-Sep-2022, Amendment No. 03 dated 12-Oct-2022, Amendment No. 04 dated 15-Oct-2022, Amendment No. 05 dated 25-Oct-2022, Amendment No. 06 dated 28-Oct-2022, Amendment No. 07 dated 04-Nov-2022 & Amendment No. 08 dated 18-Nov-2022.

Following changes are being incorporated in the tender:

**SECTION VII**  
**TECHNICAL SPECIFICATIONS**

Sch. No 6. Continuous Renal Replacement Therapy (CRRT)		
Tender spec Para	TENDER SPECIFICATION	READ AS
3.11	Should have Programmable Substitution solution flow rate: 100-8000 mL/Hr	Should have Programmable Substitution solution flow rate: <b>600-8000 mL/Hr</b>
3.12	Should have Dialysate Flow rate: <b>100-8,000 mL/Hr or better</b>	Should have Dialysate Flow rate: <b>600-4800 mL/Hr or better</b>

**Note:**

- i. All other contents of the tender enquiry including terms & conditions remain unaltered.
- ii. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.
- iii. EMD should be valid up to 225 days or more from the date of submission mentioned in TED & its subsequent amendments.
- iv. **For Tender ID 2022\_HLL\_125318\_6 (Continuous Renal Replacement Therapy) only – Techno Commercial Compliance Sheet v1.1 is being uploaded on CPP Portal. The same need to be downloaded for submission. It will supersede any other uploaded earlier.**