

**Amendment no. 10**

09.06.2021

**Sub: Amendment to the referred tender enquiry**

**Ref.: Tender Enquiry HITES/PCD/AIIMS-IV/30/Ortho/19-20 Dated 15-02-2020 read with amendment no 1 dated 20.02.2020 and Notice on Extension of Due Dates dated 25.03.2020, amendment no 2, 3, 4, 5, 6, 7, 8 & 9 dated 03.03.2021, 03.03.2021, 24.03.2021, 07.04.2021, 20.04.2021, 03.05.2021, 19.05.2021 & 25.05.2021 respectively**

The following changes are being incorporated in the following mentioned Tender ID only

- 2021\_HLL\_72064\_4 (Suction Machine)

**Section VII**  
**Technical Specification**

**Item No. 4 Suction Machine (High Vacuum)**

<b>Page &amp; Para no.</b>	<b>Technical Specification</b>	<b>Read as</b>
Page 52; Para 8	One plastic suction jar cover, steam sterilizable to be provided extra.	One suction jar cover, steam sterilizable to be provided extra.
Page 52; Para 9	Two extra suction jar (Plastic) of capacity 4-5 ltrs. Should be quoted along with accessories like lid, tubing etc. with the equipment to make the unit functional.	Two extra suction jar <b>(Polysulfone/Polycarbonate)</b> of capacity 4-5 ltrs. Should be quoted along with accessories like lid, tubing etc. with the equipment to make the unit functional.

**All other contents of the Tender enquiry including terms & conditions remain unaltered.**

**Note:**

- I. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**
- II. Technical Compliance Sheet 'Ver 1.1' and Price bid Sheet 'Ver 1.1' in excel form against referred tender are being uploaded under 'Corrigendum' on CPPP. The same need to be downloaded for submission**