

## Amendment no. 5

**Sub: Technical Specification Amendment for the tender of Onco-Anesthesia and Palliative Medicine for NCI - AIIMS (Jhajjar) , New Delhi-110029.**

**Ref.: GeM Bid No GEM/2023/B/3020284 Dated: 23-01-2023 for item 'Video Laryngoscope (Qty: 13 nos.)'**

The following changes are being incorporated in the above referred GeM Bid Number only.

Sr. No.	Tender Specification Page & Para	TENDER SPECIFICATION	Amendment
3	PAGE NO 1 OF ATC	THE MONITOR SHOULD BE TOUCH SCREEN & SHOULD HAVE 02 PARTS TO CONNECT TWO VIDEO LARYNGOSCOPES AT ONE TIME & PICTURE CAN BE SWAPPED USING TOUCH SCREEN.	Amended as:- THE MONITOR SHOULD HAVE 02 PARTS TO CONNECT TWO VIDEO <b>LARYNGOSCOPES/VIDEO ENDOSCOPE (EITHER OF ANY TWO)</b> AT ONE TIME & PICTURE CAN BE SWAPPED USING TOUCH SCREEN.
4	Page 1	Monitor resolution should be minimum 1920 X 1200 pixels in 16:9 format.	Amended as:-Monitor resolution should be minimum <b>1024 X 768 pixels.</b>
9	Page 1	The quoted model should be European CE or USFDA approved	Amended as :- quoted model should be European CE or USFDA approved or BIS or EC declaration of conformity along with Iso 13485
11	Page 1	Monitor should have two ports to connect two video laryngoscope blades at one time and picture can be swapped using touch screen.	Amended as:-THE MONITOR SHOULD HAVE 02 PARTS TO CONNECT <b>TWO VIDEO LARYNGOSCOPES/VIDEO ENDOSCOPE (EITHER OF ANY TWO)</b> AT ONE TIME & PICTURE CAN BE SWAPPED USING TOUCH SCREEN.
14	Page 1	Soft bag from same manufacturer should be supplied to place the monitor and system can also be operated without taking monitor out from the bag.	<b>Soft bag/Hard plastic</b> from same manufacturer should be supplied to place the monitor and system can also be operated without taking monitor out from the bag.
15	Page 1	Adult and Pediatric angulated Magill forceps from same manufacturer to be provided for foreign body removal and for assisting nasal intubation while using blades	Adult and Pediatric angulated Magill forceps from <b>same/reputed manufacturer and compatible with the supplied scope</b> to be provided for foreign body removal and for assisting nasal intubation while using blades
16	Page 1	<ul style="list-style-type: none"> <li>Blades should have anti fogging mechanism.</li> </ul>	Amendeed as:- <b>DELETED</b>

26	Page 1	Stand to hang monitor with tray for laryngoscope blades should be provided from same manufacturer with height 120 cm, rollable with five legs and antistatic castors, crossbar 25 cm x diameter 25 mm for positioning the monitor, with tray, dimensions (w x d x h): 30 x 20 x 10 cm.	Amended as:-Stand to hang monitor with tray for laryngoscope blades should be provided from same manufacturer with height approx. 120 cm, rollable with five legs and antistatic castors, crossbar 25 cm x diameter 25 mm for positioning the monitor, with tray, dimensions <b>approx</b> (w x d x h): 30 x 20 x 10 cm.
----	--------	--	--

**Important Note:**

1. Prospective Bidders are advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid.
2. The Bid Security/EMD shall be valid for 110 days from Techno – Commercial Tender opening date. The techno-commercial tender opening date shall be considered as mentioned in bid document read with it's amendment(s), if any.
3. In reference to the extension of bid opening date as mention on GeM portal, Participating bidders, who are submitting their Bid security/EMD in the form of Bank Guarantee (BG) or FDR are instructed to extend the validity of their BG/FDR accordingly, i.e. required validity of EMD is to be kept in line with the extended bid opening date.
4. Traders/resellers/distributors/authorized agents will not be considered for availing benefits under PP Policy 2012 for MSEs as per MSE guidelines issued by MoMSME
5. Bidders shall ensure that their tender(s), complete in all respects, are submitted online through **GeM portal only**.