

**Amendment No. 3****05.12.2022****Sub: Amendment to the Global Tender Enquiry Document****Ref.: Notice Inviting Bid ref. HITES/PCD/MCH-AIIMS/07/22-23 dated 28.09.2022 read with its Amendment no. 01 & 02 dated 20.10.2022 & 17.11.2022 and Corrigendum 1 for Auto Bid Extension.**

The following changes have been authorised and are being incorporated in the above referred Bidding Document.

**SECTION – I****NOTICE INVITING BIDS (NIB)**

<b>Description</b>	<b>Existing</b>	<b>Amended as</b>
Bid Submission End Date & Time	06-12-2022 at 02.00 pm	<b>20-12-2022</b> at 02.00 pm
Closing date & time for submission of tender fee and <b>EMD in physical form</b>  Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender fee and EMD within the mentioned date and time	07-12-2022 at 02.00 pm	<b>21-12-2022</b> at 02.00 pm
Bid Opening Date & Time	07-12-2022 at 02.30 pm	<b>21-12-2022</b> at 02.30 pm

**SECTION – VII****TECHNICAL SPECIFICATION AND GENERAL POINTS****A. TECHNICAL SPECIFICATION:****Item No. 1 (Tender ID: 2022\_HLL\_130264\_1)****Grossing Stations**

<b>Sr. No.</b>	<b>Tender Page &amp; Para</b>	<b>Existing Tender Specification</b>	<b>Amended as</b>
1	Pg. 45 Para 4 Dimensions	Length: minimum 3 1/2 feet Depth: minimum 2 1/2 feet Height: minimum 3 feet	Length: <b>5 feet</b> Depth: <b>3 feet</b> Height: <b>5-7 feet</b>

<b>Sr. No.</b>	<b>Tender Page &amp; Para</b>	<b>Existing Tender Specification</b>	<b>Amended as</b>
2	Pg. 45 Para 9	Sink: Removable seamless deep sink with hot/cold water facility for plugging in of additional appliances on work surface.	Sink: <b>Removable/fixed</b> seamless deep sink with hot/cold water facility for plugging in of additional appliances on work surface
3	Pg. 46 Para 22	Programmable microprocessor allows clock time to be changed and the filter replacement alarm to be reset.	<b>Deleted</b>

**Important Note to Prospective Bidders:**

- (i) Based on the above amendment, Technical Compliance Sheet 'TC\_2022\_HLL\_130264\_1\_Ver1.0' in CPP portal has been superseded by a new version i.e. 'TC\_2022\_HLL\_130264\_1\_Ver1.1'. The same need to be downloaded for submission by participating bidders.
- (ii) In reference to the extension of bid opening date, participating bidders, who had submitted or are submitting their Bid security/EMD in the form of Bank Guarantee (BG) or FDR are instructed to extend the validity of their BG/FDR accordingly, i.e. required validity of EMD is to be kept in line with the extended bid opening date.

**All other contents of the Global Tender Enquiry Document including terms & conditions remain unaltered.**