

Response To Pre-Bid Queries (Pre-Bid date: 30.01.2023)

Tender Ref:- GEM/2023/B/3031041

Item Name: THROMBOELASTOGRAM

Sr. No.	Tender Specification Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	TSEC Recommendations
1	Page 1, Point 1	The equipment should be a modular system with the facility to upgrade from one to 4 analyzing channels with in the main machine and should be able to run all the analyzing channels simultaneously without use of any a-D (Analog to Digital)	Measures same Viscoelastic properties as previous generation of Thromboelastography	No change considered
	Page 1, Point 2	Equipment must have complete protection mechanism for Torsion sensor by integrating automatic electromagnetic breaks and support system.	Equipment must have complete protection mechanism for Torsion sensor by integrating automatic electromagnetic breaks and support system. OR Automatically Loaded micro fluidic Cartridges	No change considered
	Page 1, Point 3	The complete procedure starting from loading of sample, heating of blood sample to mechanical synchronization with electronic components should be in one step and fully automatic without any step being done manually or considered.	Equipment should be compact, simple to use and does not require dedicated staff / Technician	No change considered
	Page 1, Point 6	Such equipments which have the technology for consumables to be reused multiple times without incurring any additional cost will be considered.	Delete. According to FDA norms consumables used once, cannot be repeated twice for different patients' blood sample.	No change considered
	Page 2/ 7	MII Purchase Preference - Yes	MII Purchase Preference – No	No change considered