

Response To Pre-Bid Queries (Pre-Bid date: 30.01.2023)				
Tender Ref:- GEM/2023/B/3020284				
Item Name: Video Laryngoscope				
Sr. No.	Tender Specification Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	TSEC Recommendations
1	Page 4	Minimum 50% and 25% Local Content required for qualifying as Class 1 and Class 2 Local Supplier	1A. SUBMISSION OF EMD AS PER GFR 2017 RULE NO 170, WE REQUEST YOU TO ALLOW THE BIDDER TO UPLOAD BID SECURITY DECLARATION DOCUMENT ON RS. 100/- STAMP PAPER IN PLACE OF EMD.	No change considered
2	NA	NA	1B. MII CONTENT THE MII CONTENT CONDITION SHOULD BE DELETED BECAUSE THIS IS NOT MANUFACTURED IN INDIA & SECONDALY ALL THE CERTIFICATES i.e., EUROPEAN CE & USFDA BECOME NULL & VOID WHEN THE INDIAN COMPANY ADDS ANY MII CONTENT.	No change considered
3	PAGE NO 1 OF ATC	THE MONITOR SHOULD BE TOUCH SCREEN & SHOULD HAVE 02 PARTS TO CONNECT TWO VIDEO LARYNGOSCOPES AT ONE TIME & PICTURE CAN BE SWAPPED USING TOUCH SCREEN.	THE MONITOR SCREEN SHOULD BE NON-TOUCH SCREEN & SHOULD HAVE ONE PORT TO CONNECT ONE VIDEO LARYNGOSCOPE. SPARE MONITOR SHOULD BE ALLOWED FOR 2ND VIDEO LARYNGOSCOPE	Amended as:- THE MONITOR SHOULD HAVE 02 PARTS TO CONNECT TWO VIDEO LARYNGOSCOPES/VIDEO ENDOSCOPE (EITHER OF ANY TWO) AT ONE TIME & PICTURE CAN BE SWAPPED USING TOUCH SCREEN.
4	Page 1	Monitor resolution should be minimum 1920 X 1200 pixels in 16:9 format.	Monitor resolution should be minimum 1024 X 768 pixels	Amended as:-Monitor resolution should be minimum 1024 X 768 pixels.
			THE NORMAL RESOLUTION OF MONITOR FOR VIDEO LARYNGOSCOPE IS 1024X768 RCB & SO IT SHOULD BE CHANGED TO HD 1024X768 RCB.	
5	Page 1	Monitor should have Picture-in-Picture & side-by-side mode to view images from 2 different blades or flexible video scopes.	SHOULD BE CHANGED TO 02 SEPARATE MONITOR FOR 02 NO OF LARYNGOSCOPE BLADES OR TWO FLEXIBLE SCOPES.	No change considered

6	Page 1	2D. BLADES & CABLES SHOULD BE TRULLY IMMERSIBLE IN DISINFECTING SOLUTION.	IT SHOULD BE CHANGED TO BLADES SHOULD BE FULLY IMMERSIBLE IN DISINFECTING SOLUTION FOR COMPLETE DISINFECTION & ALL THE REUSABLE BLADES SHOULD BE IPX8 COMPLIANT.	No change considered
7	Page 1	Blade size: 2, 3, 4 and Difficult airway blade (1 no. Each)	3D. THE SIZE OF THE BLADES SHOULD BE CHANGED TO OR SHOULD BE SELECTABLE FROM THE FOLLOWING AVAILABLE SIZES MAC CURVED ADULT – 1,2,3,4,5 MILLER STRAIGHT NEONATAL / PAED. – 00,0,1 MAC D HYPER ANGULATED BLADES FOR DIFFICULT INTUBATION – 1,2,3,4,5 MAC 0 & MILLER 2 ST. STEEL REUSABLE BLADES ARE NOT AVAILABLE IN THE MARKET TO THE BEST OF OUR KNOWLEDGE.	No change considered
8	Page 1	Blades can be sterilized using Sterrad and Steris system. Thermal disinfection upto 93 degrees and Chemo-thermal disinfection upto 65 degrees should be permissible.	THE SPECIAL ANGULATED BLADES FOR DIFFICULT INTUBATION SHOULD HAVE MORE THAN 65 DEGREE ANGLE OF VIEW.	No change considered
9	Page 1	The quoted model should be European CE or USFDA approved	THE EUROPEAN CE SHOULD BE CHANGED TO EC DECLARATION OF CONFIRMITY AS THIS FALLS UNDER THE CATEGORY OF MEDICAL DEVICE CLASS 1.	Amended as :- quoted model should be European CE or USFDA approved or BIS or EC decleration of conformity along with Iso 13485
10	Page 1	Screen 8 to 12 inch in size for display with touch screen to control features with HDMI output for connecting to a big screen which can display picture simultaneously on both screens.	Screen 3.5 inch in size for display with touch screen to control features with HDMI output for connecting to a big screen which can display picture simultaneously on both screens	No change considered
11	Page 1	Monitor should have two ports to connect two video laryngoscope blades at one time and picture can be swapped using touch screen.	Monitor should have Single ports to connect two video laryngoscope blades at one time and picture can be swapped using touch screen	Amended as:-THE MONITOR SHOULD HAVE 02 PARTS TO CONNECT TWO VIDEO LARYNGOSCOPES/VIDEO ENDOSCOPE (EITHER OF ANY TWO) AT ONE TIME & PICTURE CAN BE SWAPPED USING TOUCH SCREEN.
12	Page 1	Monitor Should have a facility to connect flexible scope directly without any special coupler or accessory.	This point to kindly deleted, it's a one company Specific	No change considered

13	Page 1	Monitor should have Picture-in-Picture & side-by-side mode to view images from 2 different blades or flexible video scopes.	This point to kindly deleted, it's a one company Specific	No change considered
14	Page 1	Soft bag from same manufacturer should be supplied to place the monitor and system can also be operated without taking monitor out from the bag.	Soft / hard plastic bag from same Manufacturer should be supplied to place the monitor and system can also be operated without taking monitor out from the bag	Soft bag/Hard plastic from same manufacturer should be supplied to place the monitor and system can also be operated without taking monitor out from the bag.
15	Page 1	Adult and Pediatric angulated Magill forceps from same manufacturer to be provided for foreign body removal and for assisting nasal intubation while using blades	Adult and paediatric Magill forceps / from same manufacturer to be provided for foreign body removal and for assisting nasal intubation while using blades	Adult and Pediatric angulated Magill forceps from same/reputed manufacturer and compatible with the supplied scope to be provided for foreign body removal and for assisting nasal intubation while using blades
16	Page 1	<ul style="list-style-type: none"> Blades should have anti fogging mechanism. 	This point to kindly deleted, it's a one company Specific	Amendeed as:- DELETED
17	Page 1	<ul style="list-style-type: none"> Macintosh blade size 0 producing resolution of minimum 800p. 	This point to kindly deleted, it's a one company Specific	No change considered
18	Page 1	<ul style="list-style-type: none"> Macintosh blade size 2 producing resolution of minimum 800p. 	This point to kindly deleted, it's a one company Specific	No change considered
19	Page 1	<ul style="list-style-type: none"> Macintosh blade size 3 producing resolution of minimum 800p. 	This point to kindly deleted, it's a one company Specific	No change considered
20	Page 1	<ul style="list-style-type: none"> Macintosh blade size 4 producing resolution of minimum 800p. 	This point to kindly deleted, it's a one company Specific	No change considered
21	Page 1	<ul style="list-style-type: none"> Miller blade size 0 producing resolution of minimum 800p. 	This point to kindly deleted, it's a one company Specific	No change considered
22	Page 1	<ul style="list-style-type: none"> Miller blade size 1 producing resolution of minimum 800p. 	This point to kindly deleted, it's a one company Specific	No change considered
23	Page 1	<ul style="list-style-type: none"> Miller blade size 2 producing resolution of minimum 800p. 	This point to kindly deleted, it's a one company Specific	No change considered

24	Page 1	<ul style="list-style-type: none"> Special angulated Adult Blade for difficult intubation with device for introduction of suction catheter of size 16-18 Fr., angle of view should be 80 degree or more should be provided with stylet from same manufacturer. 	This point to kindly deleted, it's a one company Specific	No change considered
25	Page 1	<ul style="list-style-type: none"> Special angulated Pediatric Blade for difficult intubation with device for introduction of suction catheter of size 16-18 Fr., angle of view should be 80 degree 	This point to kindly deleted, it's a one company Specific	No change considered
26	Page 1	Stand to hang monitor with tray for laryngoscope blades should be provided from same manufacturer with height 120 cm, rollable with five legs and antistatic castors, crossbar 25 cm x diameter 25 mm for positioning the monitor, with tray, dimensions (w x d x h): 30 x 20 x 10 cm.	This point to kindly deleted, it's a one company Specific	Amended as:-Stand to hang monitor with tray for laryngoscope blades should be provided from same manufacturer with height approx. 120 cm, rollable with five legs and antistatic castors, crossbar 25 cm x diameter 25 mm for positioning the monitor, with tray, dimensions approx (w x d x h): 30 x 20 x 10 cm.
27	Page 2	BOQ		No change considered
28	Page 2	Blade size: 2, 3, 4 and Difficult airway blade (1 no. Each)	This point to kindly deleted, it's a one company Specific	No change considered