

28.03.2022

Amendment No. 13**Sub: Extension of due dates to the referred tender enquiry****Ref.: HITES/PCD/AIIMS-IV/42/Rad-Onc/21-22 dated 13.09.2021 read with its amendment no 1, 2, 3, 4, 5, 6, 7, 8, 9, 10, 11 and 12 dated 22.10.2021, 08.11.2021, 26.11.2021, 20.12.2021, 27.12.2021, 04.01.2022, 12.01.2022, 18.01.2022, 01.02.2022, 16.02.2022, 28.02.2022 and 14.03.2022**

The following changes are being incorporated in the above referred Tender Enquiry Document.

**Section – VII
Technical Specifications**

Tender ref: HITES/PCD/AIIMS-IV/42/Rad-Onc/21-22 dated 13.09.2021		
Item no 03. CT Simulator System (Tender ID: 2021_HLL_88075_3)		
Technical Para No.	Existing Technical Specification	Amended As
10.10	High Contrast Spatial Resolution: Specify the high contrast resolution. Clearly specify the phantom used, scan time, mA, filter for image reconstruction, scan field, dose and MTF	High Contrast Spatial Resolution: It should be 15 lines pair per cm or better (for 50 cm FOV) maximum at 0% / 2% MTF. Clearly specify the phantom used, scan time, mA, filter for image reconstruction, scan field, dose and MTF.
Added Para	Added Para: The following to be provided : a. Laser check tool:- QA tool to verify iso-centre matching with laser system	Deleted
Added Para	Added Para: The following to be provided : b. Contrast Injector	Added Para : 1 Number of Dual head Contrast Injector to be supplied.
Added Para	Added Para: The following to be provided : c. Optional items :- 1. Portal Dosimetry for quick treatment verification for complex treatment. 2. 4D Gating with breath hold treatment option 3. 6D Couch	Deleted

All other contents of the Tender enquiry including terms & conditions remain unaltered.**Note:**

- I. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**