

Response To Pre-Bid Queries (Pre-Bid date: 17.08.2022)				
Bid Number: GEM/2022/B/2395412 Dated: 29-07-2022				
Item Name: Titanium Bone Plating Kit for Oral and Maxiofacial Surgery				
Sr. No.	Buyer Specification Document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	Reply to Pre-bid Queries
1	Pg 1 Para 8	All Instruments / implants should be CE / US FDA, should have certification	The Instrument should be USFDA or European CE certified and ISO.	Added Para:- Instrument should be USFDA or European CE or BIS or ISO 13485 certified
2	NA	Suggested by Bidder	Bidder should quote for all the instruments of the set . Atleast 75% of the quoted instruments should be from same manufacturer and from another manufacturer 25%	Added Para:-Bidder should quote for all the instruments of the set . Atleast 75% of the quoted instruments should be from same manufacturer and from another manufacturer 25%