

Response To Pre-Bid Queries (Pre-Bid date: 17.08.2022)				
Bid Number: GEM/2022/B/2392150 Dated: 28-07-2022				
Item Name: Vascular Set				
Sr. No.	Buyer Specification Document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	Reply to Pre-bid Queries
1	NA	Suggested by Bidder	The instruments should be CE or BIS certified or ISO 13485 certified	No Change considered
2	NA	Suggested by Bidder	Should be US FDA / European CE (From 4 Digit Notified Bodies under Directive: 93/42/EEC Medical devices). Reasoning: For assurance of Instrument quality as per international standard.	<b>Added Para:-</b> Instrument should be USFDA or European CE or BIS or ISO 13485 certified
3	NA	Suggested by Bidder	The Instrument should be USFDA or European CE certified and ISO.	
4	NA	Suggested by Bidder	Bidder should quote for all the instruments of the set . Atleast 75% of the quoted instruments should be from same manufacturer and from another manufacturer 25%	<b>Added Para:-</b> Bidder should quote for all the instruments of the set . Atleast 75% of the quoted instruments should be from same manufacturer and from another manufacturer 25%
5	NA	Suggested by Bidder	Vascular Set Please provide the all sizes.	No Change considered
6	NA	Suggested by Bidder	Eyelid Retractor: Please provide the all sizes.	Clarified as:- 6-8"