

Response To Pre-Bid Queries (Pre-Bid date: 17.08.2022)				
Bid Number:GEM/2022/B/2395324 Dated: 29-07-2022				
Item Name :Multipurpose Thompson retractor set				
Sr. No.	Buyer Specification Document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	Reply to Pre-bid Queries
1	NA	Suggested by Bidder	As the spec mentions a particular brand "Thompson" and the specs are also of the single brand we request to please amend spec to permit other high quality brands to also participate giving equal platform to participants	May be Amended as:- Multipurpose self retracting retractor set
2	Pg 1 Para 3	Should have various detachable blade options in different configurations for retraction of abdominal wall, solid and hollow viscera	Should have various detachable or Fixed blade options in different configurations for retraction of abdominal wall, solid and hollow viscera	No change considered
3	3	All blades should be detachable from blade holder & simultaneously different blade should be interchangeable to any / all blade holders.	All blades should be detachable or alternate fixed option from blade holder . Incase of Detachable option different blade should be interchangeable to any / all blade holders. we would request that all term may be deleted,and "option" term may be added as blades shouldhave option to be detachable.	No change considered
4	Pg 1 Para 5	Blade holder should have inbuilt angling system on the blade holder.	Blade holder should have inbuilt angling system on the blade holder if applicable. we would request that this may be amended asangling/ rotatable system	May be Amended as:- Blade holder should have inbuilt angling system on the blade holder if applicable.
5	Pg 1 Para 5	Blades & blade holder should have fixed retraction features to hold organ.	Blades & blade holder should have fixed retraction features to hold organ.	No change considered
6	Pg 1 Para 6	Blade holder rods should have ratchet mechanism in built for retraction.	Blade holder rods should have ratchet or quick lock blade mechanism in built for retraction	May be Amended as:- Blade holder rods should have ratchet or quick lock blade mechanism in built for retraction
7	Pg 1 Para 11	Equipment must be US – FDA approved or European CE with 4 digit notified body number	The Instrument should be USFDA or European CE certified and ISO.	May be Amended as:- Instrument should be USFDA or European CE or BIS or ISO 13485 certified
8	Pg 1 Para 1	Elite II Rail Clamp w/2 Cam Joints 18"	2 Elite II Rail Clamp w/2 Cam Joints 18" Or 18" Field Post with 2 Couplers Main Holder Table post	No change considered No change considered
9	Pg 1 Para 2	Crossbar w/2 Cam Joints 24"	1 CamJoint or Coupler for Bilateral Crossbar Crossbar -1	No change considered No change considered
10	Pg 1	34" Bilateral Crossbar Hinged	1 Bilateral Cross Bar, Hinged 11½x8½x11½ or	No change considered

	Para 3	11½x8½x11½		Main Retainer -1	No change considered
11	Pg 1 Para 4	24" Angled Arm 11" x 13"	2	Angled Arms Large	May be Amended as:- Angled Arms Large
12	Pg 1 Para 5	24" Angled Arm 8" x 16"		Kindly amend to read as " 24" Angled Arm Angled Arm 8" x 16" or 24" Field Post "	No change considered
			2	Angled Arms Medium	No change considered
13	Pg 1 Para 6	18" Angled Arm 8" x 10" @45°	2	18" Angled Arm 8" x 10" @45° or Field Post 12"	No change considered
				Angled Arms Small	No change considered
14	Pg 1 Para 7	Ring Half Hinged 14½"	2	Ring Half Hinged 14½" or Square Retractor frame Reusable Hinged Single Arm	May be Amended as:- Ring Half Hinged 14½" or Square Retractor frame Reusable
15	Pg 1 Para 8	Ring Extension Hinged 3"	2	We would request that this may be deleted as this is only required for full ring, but full ring has not been asked as mandatory, you have given option at above point at 2 to quote any of three options available ring / half ring / single arm configuration. Something which has not been asked as mandatory may be made optional or else options for others also may also be provided like single arm Ring Extension Hinged 3" or Extension Stays	No change considered
16	Pg 1 Para 9	Extension Arm 14" w/Cam Latch Mount	2	Extension Extension Arm 14" w/Cam Latch Mount OR Bridge Kit Set	May be Amended as:- Extension Arm 14" w/Cam Latch Mount OR Bridge Kit Set
17	Pg 1 Para 10	S-Lock Handle Angling 8"	2	Kindly amend the same as this is proprietary to a single company and is not applicable for all brands. Hence kindly amend to read as "S-Lock Handle Angling 8" - 2nos or Quick lock clamp 2 Nos" Angling Blade Holder	May be Amended as:- S-Lock Handle Angling 8" - 2nos or Quick lock clamp 2 Nos"
18	Pg 1 Para 11	S-Lock Micro-Adj II Clip-on Angling 7¾"	2	Kindly amend the same as this is proprietary to a single company and is not applicable for all brands. Hence kindly amend to read as "S-Lock Micro-Adj II Clip-On Angling 7¾" or Quick lock clamp 2nos " Blade Holder	May be Amended as:- S-Lock Micro-Adj II Clip-On Angling 7¾" or Quick lock clamp 2nos "
19	Pg 1 Para 12	S-Lock Micro-Adj II Clip-On Angling 10"	4	Kindly amend the same as this is proprietary to a single company and is not applicable for all brands. Hence kindly amend to read as "S-Lock Micro-Adj II Clip-On Angling 10" or Quick Lock blade Clamps 4 nos" Clamps	May be Amended as:- S-Lock Micro-Adj II Clip-On Angling 10" or Quick Lock blade Clamps 4 nos"
20	Pg 1 Para 13	S-Lock Handle Held Adapter	1	This is patented to Thompson Surgical; we would request that option against this may be provided as we have complete set available which has been asked at point 10, so we don't need separate adapter as asked at 13, so this point may be made optional.	May be Amended as:- S-Lock Handle Held Adapter or Quick Lock blade Clamps 1 nos"

				S-Lock Handle Held Adapter or Quick Lock blade Clamps 1 nos"	
21	Pg 1 Para 14	Angling Head Wrench	1	Kindly amend the same as this is proprietary to a single company and is not applicable for all brands. Hence kindly amend to read as "Angling Head Wrench if applicable Angling head	May be Amended as:- Angling Head Wrench if applicable
22	Pg 1 Para 15	Instrument Case 26 x 10 x 5"	2	Please confirm you need from same make as manufacturer or local make can also be supplied Instrument Case	May be Amended as:-Instrument Case
23	Pg 1 Para 16	Balfour w/o Lips 70mm x76mm(2¾" x 3")-SL	2	Abdominal blades 78x65 mm (+/- 10% variation)	May be Amended as:-10% variation in sizes is acceptable
24	Pg 1 Para 17	Balfour w/o Lips 83mm x76mm (3¼" x3")-SL	2	Abdominal blades 80x90 mm (+/- 10% variation)	
25	Pg 1 Para 18	Balfour 83mm x 57mm (3¼" x 2¼")-SL	1	Abdominal blades 60x85 mm (+/- 10% variation)	
26	Pg 1 Para 19	Fence 102mm x 127mm (4" x 5")-SL	1	Intestinal Blade 100x120mm	
27	Pg 1 Para 20	Kelly 64mm x 76mm (2½" x 3")-SL	1	Kelly 64x80	
28	Pg 1 Para 21	Kelly 76mm x 89mm (3" x 3½")-SL	1	Kelly 72x82	
29	Pg 1 Para 22	Soft Malleable 25mm x 102mm (1" x 4")-SL	1	Malleable 25x100	
30	Pg 1 Para 23	Malleable 51mm x 203mm (2" x 8") - SL	1	Malleable 50x200	
31	Pg 1 Para 24	St Marks 64mm x 178mm (2½"x7") w/Lip-SL	1	St Marks 60x174	
32	Pg 1 Para 25	Harrington 64mm x 152mm (2½" x 6") SL	1	Harrington 65 x 175	
33	Pg 1 Para 26	Deaver 51mm x 127mm (2" x 5") - SL	1	Deaver 50x125	
34	Pg 2 Para 27	Richardson 51mm x 127mm (2" x 5")-SL	1	Richardson 38x37	May be Deleted
35	Pg 2 Para 28	Radio Richardson 51mm x 178mm - SL	1	Richardson 51mm x 178mm We would request that this may be made optional as in open surgery C arm is not used, so radiolucent blade as asked is not required and same is proprietary to Thompson surgical. As at point 27, required blade is already asked. So we would request that this point may be made optional.	
36	Pg 2 Para 29	Malleable Finger 6" - SL	1	Malleable Finger -1	No change considered
37	Pg 2 Para 30	Y Bladder - SL	1	Bladder Retractor - 1	No change considered
38	NA	Suggested by Bidder	Bidder should quote for all the instruments of the set . Atleast 75% of the quoted instruments should be from same manufacturer and from another manufacturer 25%		Added Para:-Bidder should quote for all the instruments of the set . Atleast 75% of the quoted instruments should be from same manufacturer and from another manufacturer 25%