

Response To Pre-Bid Queries (Pre-Bid date: 17.08.2022)				
Bid Number: GEM/2022/B/2388362 Dated: 27-07-2022				
Item Name:Rigid Nasal Endoscope 4mm; 0 Degree, 30 & 45 degree				
Sr. No.	Buyer Specification document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	Reply to Pre-bid Queries
1	Pg 1 Para 4 (Under Nasal Endoscope 0deg 4mm)	Rod lenses system, Length: 18-19 cms	We request to Kindly amend it as "Insertion tube diameter: 4.5-5.9 mm"	No change considered
2	Pg 1 Para 4 (Under Nasal endoscope 30° 4.0MM)	Rod lenses system, Length: 18-19 cms	We request to kindly amend it as " bridge for irrigation with two lateral irrigation ports and working channels "	No change considered
3	Pg 1 Para 4 (Under Nasal endoscope 45° 4.0MM)	Rod lenses system, Length: 18-19 cms	We request to kindly amend it as "Rod lenses system, length 16-19cms"	May be amended as:- Rod lenses system, Length: 16-19 cms
4	NA	Suggested by Bidder	Incorporation of safety certification i.e. USFDA or EUROPEAN CE by 4 Digit Notified Body.	may be amended as:- Manufacturer should have ISO 13485 certification
5	NA	Suggested by Bidder	Permissible variations in size length of instruments.	May be Amended as:-10% variation in sizes is acceptable
6	NA	Suggested by Bidder	Please allow us to quote 90% instruments from the same manufacturer.	Added Para:-Bidder should quote for all the instruments of the set . Atleast 75% of the quoted instruments should be from same manufacturer and 25% can be from different manufacturer

Commercial Query

Sr. No.	Buyer Specification document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	Reply to Pre-bid Queries
1	NA	Suggested by Bidder	Kindly make amendment in the MII clause which will enable us to quote quality products as per the tender requirements which are not made in India.	No change considered
2	NA	Suggested by Bidder	Allow us to quote substitute products irrespective of the item/name referring to a particular manufacturer/proprietor where the functionality and quality of the outcomes will not be compromised while keeping fare competition	No change considered
3	NA	Suggested by bidder	We request for MII (Make In India) to be exempted for this tender and should be floated as global tender.	No change considered