

Response To Pre-Bid Queries (Pre-Bid date: 17.08.2022)				
Bid Number: GEM/2022/B/2387470 Dated: 27-07-2022				
Item Name: Complete Lower Tract Endourology (Adult Cystoscope & Resectoscope)				
Sr. No.	Buyer Specification document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	Reply to Pre-bid Queries
1	Pg 1 Para 1	Straight Forward 0-degree Telescope, diameter-4mm, Autoclavable, Length 30cm, fiber optic light transmission incorporated	We request to kindly amend it as "4mm 0 / 12-degree Telescope having working length 28cm or more" Remarks: Olympus Optical Urethrotome is compatible with 12-degree telescope only	no change considered
2	Pg 1 Para 2	Forward oblique 30-degree Telescope, diameter 4mm, Autoclavable, Length 30cm, fiber optic light transmission incorporated	We request to kindly amend it as "working length 28cm or more"	no change considered
3	Pg 1 Para 5	Rigid Optical Grasping forceps, double action jaws for stent removal compatible with 19/19.5Fr/ 21 Fr	We request to kindly amend it as "Rigid Optical Grasping forceps, double action jaws for stent removal compatible with 19/19.5Fr/ 21 /22.5Fr" Remarks: This as 21 Fr sheath has not been asked in serial no 3	Rigid Optical Grasping forceps, double action jaws for stent removal compatible with 19/19.5Fr/ 22/22.5Fr'
4	Pg 1 Para 7	Biopsy Forceps 7Fr each with double action jaws, flexible, length 40cm	We request to kindly amend "working length as 35-40cm"	no change considered
5	Pg 1 Para 8	Scissors of 7Fr, Single action jaws, flexible, length 40cm	We request to kindly amend "working length as 35-40cm"	no change considered
6	Pg 1 Para 10	Standard obturator & Visual obturator each with channel for flexible instruments.	We request to kindly amend it as "Standard obturator & Visual obturator each"	no change considered
7	Pg 1 Para 15	Coagulating electrode, with barrel-shaped end Kindly amend it a barrel shaped or similarmonopolar	Kindly amend it a barrel shaped or similar	no change considered

**Commercial Query**

Sr. No.	Buyer Specification document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	NA	Suggeste by bidder	We request for MII (Make In India) to be exempted for this tender and should be floated as global tender.	no change considered