

Response To Pre-Bid Queries (Pre-Bid date: 17.08.2022)				
Bid Number: GEM/2022/B/2394123 Dated: 29-07-2022				
Item Name: Abdominal/ Vaginal Hysterectomy Instrument set				
Sr. No.	Buyer Specification Document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	Reply to Pre-bid queries
1	Pg 2 Para 31	Mayo's Scissors (TC TIP)	Please confirm size and Straight/Curved	6" curved & straight each
2	Pg 2 Para 32	Kelly Clamps	Please confirm size	6"
3	Pg 2 Para 33	Right Angled Clamps	Please confirm size	6"
4	Pg 2 Para 34	Suction tips	Please confirm size	standard size
5	Pg 2 Para 35	Micro needle holder (TC TIP)	Please confirm size and Straight/Curved	6" curved & straight each
6	Pg 2 Para 36	Microscissors (TC TIP)	Please confirm size and Straight/Curved	6" curved & straight each
7	Pg 2 Para 37	Wertheim's Vaginal clamp	Please confirm size	standard size
8	Pg 2 Para 38	TC parametrium scissors	Please confirm size and Straight/Curved	6" curved & straight each
9	Pg 2 Para 41	Vascular set	Require contain in this set	Yes
10	Pg 2 Para 42	Intestinal clamps (Crushing)	Please confirm size and Straight/Curved	standard size
11	Pg 2 Para 43	Intestinal clamps (NonCrushing)	Please confirm size and Straight/Curved	standard size
12	Pg 2 Para 44	Vein loop retractor	Please confirm size	6" and 8" each
13	Pg 2 Para 42	The Instruments should be US-FDA or European CE certified.	The instruments should be CE or BIS certified or ISO 13485 certified The Instrument should be USFDA or European CE certified and ISO.	Added Para:-Manufacturer should have ISO 13485 certification
14	NA	Suggested by Bidder	Should be US FDA / European CE (From 4 Digit Notified Bodies under Directive: 93/42/EEC Medical devices). Reasoning: For assurance of Instrument quality as per international standard.	
15	NA	Suggested by Bidder	Bidder should quote for all the instruments of the set . Atleast 75% of the quoted instruments should be from same manufacturer and from another manufacturer 25%	no change considered