

| Response To Pre-Bid Queries (Pre-Bid date: 17.08.2022) | | | | |
|--|--|--|---|--|
| Bid Number: GEM/2022/B/2395587 Dated: 29-07-2022 | | | | |
| Item Name: Minor surgical set | | | | |
| Sr. No. | Buyer Specification Document Page & Para | TENDER SPECIFICATION | REPRESENTATION RECEIVED FROM THE FIRM | Reply to Pre-bid Queries |
| 1 | Pg 1 Para 2 | The Instruments should be US-FDA or European CE certified. | <p>The instruments should be CE or BIS certified or ISO 13485 certified</p> <p>The Instrument should be USFDA or European CE certified and ISO.</p> <p>Should be US FDA / European CE (From 4 Digit Notified Bodies under Directive: 93/42/EEC Medical devices).</p> <p>Reasoning: For assurance of Instrument quality as per international standard.</p> | May be Amended as:- Instrument should be USFDA or European CE or BIS or ISO 13485 certified |
| 2 | NA | Suggested by Bidder | Bidder should quote for all the instruments of the set . Atleast 75% of the quoted instruments should be from same manufacturer and from another manufacturer 25% | Added Para:- Bidder should quote for all the instruments of the set . Atleast 75% of the quoted instruments should be from same manufacturer and from another manufacturer 25% |