

Response To Pre-Bid Queries (Pre-Bid date: 17.08.2022)				
Bid Number: GEM/2022/B/2397570 Dated: 30-07-2022				
Item Name: General Surgery Set				
Sr. No.	Buyer Specification Document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	Reply to Pre-bid Queries
1	Pg1 Para 2	The Instruments should be US-FDA or European CE certified.	The instruments should be CE or BIS certified or ISO 13485 certified	May be Amended as:- Instrument should be USFDA or European CE or BIS or ISO 13485 certified
			The Instrument should be USFDA or European CE certified and ISO.	
			Should be US FDA / European CE (From 4 Digit Notified Bodies under Directive: 93/42/EEC Medical devices). Reasoning: For assurance of Instrument quality as per international standard.	
2	NA	Suggested by Bidder	Bidder should quote for all the instruments of the set . Atleast 75% of the quoted instruments should be from same manufacturer and from another manufacturer 25%	Added Para:- Bidder should quote for all the instruments of the set . Atleast 75% of the quoted instruments should be from same manufacturer and from another manufacturer 25%