

Response To Pre-Bid Queries (Pre-Bid date: 17.08.2022)				
Bid Number: GEM/2022/B/2392070 Dated: 28-07-2022				
Item Name: Dermatome set				
Sr. No.	Buyer Specification Document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	Reply to Pre-bid Queries
1	Pg 1 Para	Autoclave box (Zimmer) Containing	Autoclave box Containing Reasoning: Zimmer is Brand Name hence this word should be removed for broad based Specification Autoclave box Containing Justification: We wish to request you to kindly delete the brand name zimmer which will allow other manufacturer to participate in the tender.	May be Amended as:- Autoclave box Containing
2	Pg 1 Para 1	Dermatome handle cord: 01(one)	Dermatome handle cord Should be electrically operated. Reasoning: Only Electric Dermatome can give 4 inch skin graft and it will save wastage of skin graft.	No change cosidered
3	Pg 1 Para 2	Graft plate guard : 01(One)	Graft plate guard. The thickness control adjustment ranges from 0.75mm with increment of 0.05 mm. Resoning: As per international standard specification.	May be Amended as:- 01 (One) and <b>The thickness control adjustment ranges from 0.75mm with increment of 0.05 mm.</b>
4	Pg 1 Para 3	Screw Driver (Green colour handle): 01(One)	There should be Screw Driver Reasoning: Colour is not specific point. Justification: We wish to request you to kindly delete (Green colour handle)which allow other manufacturer to participate in the tender as colour of screw driver handle does not affect the application of the device.	May be Amended as:- Screw Driver : 01(One)
5	Pg 1 Para 14	Total Articulation should be 110-150 degree.	Blades should be supplied sterile without any need for lubrication. Reasoning: Blades is essentially require to operate Dermatome	No change cosidered
6	NA	Suggested by Bidder	The Instrument should be USFDA or European CE certified and ISO.	<b>Added para:-</b> Equipment should be ISO 13485 certified
7	NA	Suggested by Bidder	Bidder should quote for all the instruments of the set . Atleast 75% of the quoted instruments should be from same manufacturer and from another manufacturer 25%	<b>Added para:-</b> Bidder should quote for all the instruments of the set . Atleast 75% of the quoted instruments should be from same manufacturer and from another manufacturer 25%