

Response To Pre-Bid Queries (Pre-Bid date: 17.08.2022)				
Bid Number: GEM/2022/B/2397072 Dated: 30-07-2022				
Item Name: Video Transnasal Esophagoscope				
Sr. No.	Buyer Specification document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	Reply to Pre-bid Queries
1	Pg 1 Para 5	Insertion tube diameter: 4.5-5.8 mm	We request to Kindly amend it as "Insertion tube diameter: 4.5-5.9 mm"	May be amended as:- "Insertion tube diameter: 4.5-5.9 mm"

**Commercial Query**

Sr. No.	Buyer Specification document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	Reply to Pre-bid Queries
1	NA	Suggested by bidder	We request for MII (Make In India) to be exempted for this tender and should be floated as global tender.	No change considered