

16.10.2021

Amendment No. 3**Sub: Extension of due dates to the referred tender enquiry****Ref.: Tender Enquiry HITES/PCD/AIIMS-IV/39/Anst/21-22 dated 23.08.2021 Read with Amendment no 1 and 2 dated 21.09.2021 and 16.10.2021 respectively**

The following changes are being incorporated in the above referred Tender Enquiry Document.

SECTION I**NOTICE INVITING TENDER (NIT)****Existing:**

Sl.No	Tender ID	Name of the Item	Qty for 15 AIIMS	Unit	Tender processing Fee(Incl of GST@18%)	EMD
5	2021_HLL_85749_5	Fibreoptic bronchoscope with monitor and recording facility for adult and paediatric	30	Nos	2950	NIL

Read As:

Sl.No	Tender ID	Name of the Item	Qty for 15 AIIMS	Unit	Tender processing Fee (Incl of GST@18%)	EMD
5	2021_HLL_85749_5	Flexible High Definition Video Bronchoscope with Monitor and recording facility for Adult	30	Nos	2950	NIL

Added Item

Sl.No	Tender ID	Name of the Item	Qty for 15 AIIMS	Unit	Tender processing Fee(Incl of GST@18%)	EMD
18	2021_HLL_85749_18	Flexible Bronchoscope with monitor and recording facility for pediatric	15	Nos	2950	NIL

SECTION - VI
LIST OF REQUIREMENTS

Part I:**Existing**

Item No.	Tender ID	Name of the Item	Qty.	Warranty in years	CMC in years
5	2021_HLL_85749_5	Fibreoptic bronchoscope with monitor and recording facility for adult and paediatric	30	5	5

Read As:

Item No.	Tender ID	Name of the Item	Qty.	Warranty in years	CMC in years
5	2021_HLL_85749_5	Flexible High Definition Video Bronchoscope with Monitor and recording facility for Adult	30	5	5

Added item

Item No.	Tender ID	Name of the Item	Qty.	Warranty in years	CMC in years
18	2021_HLL_85749_18	Flexible Bronchoscope with monitor and recording facility for pediatric	15	5	5

Section – VII
Technical Specifications
Added Specification for Tender ID 2021_HLL_85749_18

Flexible Bronchoscope with monitor and recording facility for pediatric	
Sl. No	Technical Specification
	The Flexible Paediatric Bronchoscope should be supplied complete with light source, recording system, trolley and minimum 15" LCD/LED/TFT Monitor
A	Paediatric Scope:

1	Field of View should be 90 degree or more		
2	Depth of field should be 3 – 50 mm or better		
3	Distal end diameter should be 3.2 mm or less		
4	Insertion tube diameter should be 3.2 mm or less		
5	Channel diameter should be 1.2 mm or more		
6	Should be light weight and easy to use		
7	Working length should be 600 mm or more		
8	Can be fully immersed in disinfectant solution and water		
9	Should be European CE with 4 digit notified body number/US FDA/BIS approved OR should meet IEC 60601-1, IEC 60601-1-2 & 60601-2-18 standards and submit valid test report from any NABL accredited labs or from the labs in their country of origin (incase of foreign manufacturers) for the quoted model)		
B	Video Processor & Light source		
1	Outputs - suitable video output		
2	Unit should be compact and light weight.		
3	Light source - Combined or separate LED/150W Xenon (covered under warranty)		
4	Lamp can be turned on/off without turning off the equipment.		
5	Electronic magnification up to 1.5X by a touch of scope remote switches/adapters/coupler		
6	Should be European CE with 4 digit notified body number/US FDA/BIS approved OR should meet IEC 60601-1, IEC 60601-1-2 standards and submit valid test report from any NABL accredited labs or from the labs in their country of origin (incase of foreign manufacturers) for the quoted model)		
C	Monitor		
1	15 inches or more LCD/LED HD Monitor of Medical Grade. It should be mountable on trolley.		
D	Computer with Software		
1	Should be supplied with suitable computer system with facility for recording images and video.		
E	Trolley		
1	Suitable Trolley (closed cabinet type) to mount monitor and to store scopes, light source and all accessories		
SI NO	BOQ	Qty	UOM
1	Flexible Bronchoscope Paediatric	1	No
2	Monitor HD	1	No
3	Video Processing System with recording system	1	No

4	Light Source	1	No
5	Trolley	1	No
6	Cleaning/maintenance kit including container for diinfectant solution	1	Set
7	Foreign body forceps basket type (Optional)	2	No
8	ET tube holder	1	No
9	Leak tester	1	No

All other contents of the Tender enquiry including terms & conditions remain unaltered.

Note:

- I. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**