

Amendment no. 3

Sub: Technical Specification Amendment for the tender of Mother & Child Block -AIIMS , New Delhi-110029.

Ref.: GeM Bid No GEM/2022/B/2583967 Dated: 28-09-2022 for item 'Defibrillator with Paediatric pads with AED facility (Qty: 03 nos.)'

The following changes are being incorporated in the above referred GeM Bid Number only.

Sr. No.	ATC DOCUMENT PARA AND PAGE	EXISTING TENDER SPECIFICATION	READ AS
1	NA	NA	Added Para: Feature of Cardiac Monitor with defibrillator which should be upgradable for SpO2, NIBP, Temperature and Etco2 for wider participants.
2	Para 1 Point 7	The machine should be able to defibrillate Adult and pediatric patients. Also 2 in 1 Or separate paddles for pediatric patients should be supplied (one set with each unit).	The machine should be able to defibrillate Adult and pediatric patients. Also 2 in 1 Or separate paddles for pediatric patients should be supplied.
3	NA	NA	Added Para: Accessories should be supplied with each unit. 1. ECG cable with accessories-1 Set 2. Disposable pad- 1 No 3. ECG rolls 10 Nos 4. Adult paddles-1No 5. Paediatric Paddles- 2 Nos
7	NA	NA	Added Para:- Should have import/manufacturing license from Central licensing Authority or State licensing authority of CDSCO for Medical Devices and copy of valid license should be submitted for the quoted model. In case the vendor has not yet obtained import/manufacturing license from CDSCO for the quoted model, proof of application for CDSCO medical device license to be submitted in the bid document and valid CDSCO license to be produced at the time of supply/ NOA for the quoted model

Important Note:

1. Prospective Bidders are advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid.
2. The Bid Security/EMD shall be valid for 110 days from Techno – Commercial Tender opening date. The techno-commercial tender opening date shall be considered as mentioned in bid document read with it's amendment(s), if any.
3. In reference to the extension of bid opening date as mention on GeM portal, Participating bidders, who are submitting their Bid security/EMD in the form of Bank Guarantee (BG) or FDR are instructed to extend the validity of their BG/FDR accordingly, i.e. required validity of EMD is to be kept in line with the extended bid opening date.
4. Traders/resellers/distributors/authorized agents will not be considered for availing benefits under PP Policy 2012 for MSEs as per MSE guidelines issued by MoMSME
5. Bidders shall ensure that their tender(s), complete in all respects, are submitted online through **GeM portal only**.