

03.06.2022

**Amendment No. 11****Sub: Amendment to the referred tender enquiry**

**Ref.: .: HITES/PCD/AIIMS-IV/50/Mix/ 22-23 dated 17-03-2022 read with its Amendment No 1, 2, 3, 4, 5, 6, 7, 8, 9 and 10 dated 18.04.2022, 30.04.2022, 02.05.2022, 07.05.2022, 12.05.2022, 19.05.2022, 26.05.2022 and 31.05.2022 respectively**

The following changes are being incorporated in the above referred Tender Enquiry Document for following items only

Tender Id	Item name
2022_HLL_110121_9	Arthroscopy System

**SECTION – VII**  
**TECHNICAL SPECIFICATIONS**

Arthroscopy System (2022_HLL_110121_9)		
Tender Para	Page &	TENDER SPECIFICATION
	NA	Added Para Under '3. Powerful LED light source' : f. Lamp colour temperature : 5500K - 7500K (adjustable)
	NA	Added Para Under '3. Powerful LED light source' : g. Lamp life: Minimum of 20000 hours

All other contents of the Tender enquiry including terms & conditions remain unaltered.

Note:

- I. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids