

Amendment No. 01**Sub: Amendment to the referred tender enquiry****Ref.:** Tender Enquiry Document HITES/PCD/NIA/01/22-23 dated 05-08-2022

Following changes are being incorporated in the tender:

Section – VII
Technical Specifications

| HITES/PCD/NIA/01/22-23 dated 05-08-2022 | | |
|---|---|---|
| Item Sch. No. 10 - Fully Automated Immunoassay Analyser (2022_HLL_124548_10) | | |
| Tender Page & Para | TENDER SPECIFICATION | AMENDED AS |
| Page 62, Para 16 | Systems have the facility to test special Immunoassays parameters like Troponine T, NT-pro BNP S100 , IL 6, Vit D, P1NP, PLGF, SFLT, ACTH, anti TSHR, anti CCP, anti HCV, PAPP-A , Procalcitonine, Hepatitis B Marker assays, Human Growth Hormone, Toxo IgG & IgM, HSV-1, HSV- 2 , Rubella IgG & IgM, CMV IgG & IgM, Free Bhcg & PAPP-A, Anti Mullerian hormone (AMH), vit. D, Calcitonin, Tacrolimus, Cyclosporine, Syphilis, SARS Covid total antibody, SARS Covid IGG antibody besides the other routine immunology parameters | Systems have the facility to test special Immunoassays parameters like Troponine-T, NT-pro BNP, S100 , IL 6, Vit D, P1NP, PLGF, SFLT, ACTH, anti TSHR, anti CCP, anti HCV, Procalcitonine, Hepatitis B Marker assays, Human Growth Hormone, Toxo IgG & IgM, HSV-1 IgG, HSV- 2 IgG , Rubella IgG & IgM, CMV IgG & IgM, Free Bhcg, Anti Mullerian hormone (AMH), Calcitonin, Tacrolimus, Cyclosporine, Syphilis, SARS Covid total antibody, SARS Covid IGG antibody besides the other routine immunology parameters |
| Reagent Price List for NIA Jaipur | | |
| Page 63, Para 2 | 2. NT-pro BNP S100 - Nil | NT-pro BNP - 50 nos S100 - 50 nos |
| Page 63, Para 14 | 14. Hepatitis B Marker assays - 200 | 14. Hepatitis B Marker assays (i) Hepatitis B surface antigen (HBsAg) - 200 nos (ii) Hepatitis B surface antibody (Anti-HB) - 200 nos (iii) Hepatitis B core antibody (anti-HBc) - 200 nos (iv) IgM antibody to hepatitis B core antigen (IgM anti-HBc) - 200 nos |
| Page 63, Para 18 | 18. HSV-1 - 100 | HSV-1 IgG - 100 nos |

| HITES/PCD/NIA/01/22-23 dated 05-08-2022 | | |
|--|--------------------------|---------------------------|
| Item Sch. No. 10 - Fully Automated Immunoassay Analyser (2022_HLL_124548_10) | | |
| Page 63, Para 19 | 19. HSV- 2 - 100 | HSV-2 IgG - 100 nos |
| Page 63, Para 25 | 25. Free PAPP-A - 50 | Deleted |
| Page 63, Para 38 | 38. ESTROGEN - 700 | Estradiol-E2/E3 - 700 nos |
| Page 63, Para 40 | 40. ANDOSTENEDIONE - 700 | Deleted |
| Page 63, Para 44 | 44. SERUM IRON - 300 | Deleted |
| Page 64, Para 46 | 46. TIBC - 300 | Deleted |

Added :

Reagent Price List for Fully Automated Immunoassay Analyser , NIA Jaipur may be read as;

| Reagent Price List for NIA Jaipur | | | | | | | | | |
|---------------------------------------|--------------------|--|---------------------------|--------------------|-------------------|-----------------|--|------------------------------------|------------------------|
| PRICE QUOTE FOR IMMUNO ASSAY ANALYSER | | | | | | | | | |
| Sl. No | List of Parameters | (A) | Reagent Pack Details | | | | | | |
| | | No. of tests (approximate load over 5 years being factored for bid ranking only) | Reagent Pack - make/brand | Pack Catalogue No. | No. of tests/pack | Pack size in ml | Total No. of Reagent packs to be used for No. of tests in column "A" (i) | Unit Price per pack incl. GST (ii) | Total Price (i) x (ii) |
| 1 | Troponine T | 100 | | | | | | | |
| 2 | NT-pro BNP | 50 | | | | | | | |
| 3 | S100 | 50 | | | | | | | |
| 4 | IL 6 | 190 | | | | | | | |

| Reagent Price List for NIA Jaipur | | | | | | | | | |
|-----------------------------------|--|-------|--|--|--|--|--|--|--|
| 5 | Vit D | 500 | | | | | | | |
| 6 | P1NP | 50 | | | | | | | |
| 7 | PLGF | 50 | | | | | | | |
| 8 | SFLT | 50 | | | | | | | |
| 9 | ACTH | 400 | | | | | | | |
| 10 | anti TSHR | 200 | | | | | | | |
| 11 | anti CCP | 500 | | | | | | | |
| 12 | anti HCV | 100 | | | | | | | |
| 13 | Deleted | ----- | | | | | | | |
| 14 | Procalcitonine | 200 | | | | | | | |
| 15 | Hepatitis B Marker assays | | | | | | | | |
| (i) | Hepatitis B surface antigen (HBsAg) | 200 | | | | | | | |
| (ii) | Hepatitis B surface antibody (Anti-HB) | 200 | | | | | | | |
| (iii) | Hepatitis B core antibody (anti-HBc) | 200 | | | | | | | |
| (iv) | IgM antibody to hepatitis B core antigen (IgM anti-HBc) | 200 | | | | | | | |
| 16 | Human Growth Hormone | 300 | | | | | | | |
| 17 | Toxo IgG | 200 | | | | | | | |
| 18 | Toxo IgM | 200 | | | | | | | |
| 19 | HSV-1 IgG | 100 | | | | | | | |
| 20 | HSV- 2 IgG | 100 | | | | | | | |
| 21 | Rubella IgG | 200 | | | | | | | |
| 22 | Rubella IgM | 200 | | | | | | | |
| 23 | CMV IgG | 200 | | | | | | | |
| 24 | CMV IgM | 200 | | | | | | | |
| 25 | Free Bhcg | 50 | | | | | | | |

| Reagent Price List for NIA Jaipur | | | | | | | | | |
|---|--|-------|--|--|--|--|--|--|--|
| 26 | Deleted | ----- | | | | | | | |
| 27 | Anti Mullerian hormone (AMH) | 50 | | | | | | | |
| 28 | Calcitonin | 300 | | | | | | | |
| 29 | Tacrolimus | 50 | | | | | | | |
| 30 | Cyclosporine | 0 | | | | | | | |
| 31 | Syphilis | 500 | | | | | | | |
| 32 | SARS Covid total antibody | 400 | | | | | | | |
| 33 | SARS Covid IGG antibody | 400 | | | | | | | |
| 34 | Other routine immunology parameters (Please specify parameters separately) | 1200 | | | | | | | |
| 35 | FSH | 1200 | | | | | | | |
| 36 | PROLACTIN | 1200 | | | | | | | |
| 37 | TESTOSTERONE | 1200 | | | | | | | |
| 38 | PROGESTERONE | 700 | | | | | | | |
| 39 | Estradiol-E2/E3 | 700 | | | | | | | |
| 40 | DHEA | 700 | | | | | | | |
| 41 | Deleted | ----- | | | | | | | |
| 42 | T3 | 3600 | | | | | | | |
| 43 | T4 | 3600 | | | | | | | |
| 44 | TSH | 3600 | | | | | | | |
| 45 | Deleted | ----- | | | | | | | |
| 46 | SERUM FERRITIN | 300 | | | | | | | |
| 47 | Deleted | ----- | | | | | | | |
| 48 | SHBG | 300 | | | | | | | |
| 49 | SERUM INSULIN | 1000 | | | | | | | |
| Total price of reagents for 5 years | | | | | | | | | |
| Spares should be specified and price should be quoted separately | | | | | | | | | |

| Section II - Consumable cost (other than reagents) | | | | | | | |
|--|--|--|-------------|---------------|--|-------------------------------------|-------------------------|
| Items/ Pack Details | | | | | | | |
| Sl. No. | Type of Consumable (Bidder may add additional rows) | Details of Tests / Analyzers/ Equipment, etc. the consumable is being used for | Make/ brand | Catalogue No. | (E) Total No. of Consumable item/ packs to be used for cumulative no. of tests as detailed in column "A" of Section 'I' | Unit Price per pack incl. GST (iii) | Total Price (iii) x (E) |
| | Calibrators | | | | | | |
| | | | | | | | |
| | | | | | | | |
| | Quality controls | | | | | | |
| | | | | | | | |
| | | | | | | | |
| | Additives | | | | | | |
| | | | | | | | |
| | | | | | | | |
| | Cleaners | | | | | | |
| | | | | | | | |
| | | | | | | | |
| | Any other Consumables | | | | | | |
| | | | | | | | |
| | | | | | | | |
| | Important Note: | | | | | | |
| | Any reagent or any consumable required for performing tests, calibration, quality control, cleaning the lab system, etc. as per quantities detailed in column "A" of Section-I, if not quoted in section-I and section 'II' of APPENDIX 1 shall be provided free of cost by the bidder during the validity of the contract. | | | | | | |

| Item Sch. No. 11 - Fully Automated HbA1C/HPLC Analyser (2022_HLL_124548_11) | | |
|--|--|--|
| Tender Page & Para | TENDER SPECIFICATION | AMENDED AS |
| Page 65, Para 13 | The system should have the facility to perform either HbA1C or HbA2/Hbf without changing any column/ cartridge | The System should have the facility to perform HbA2/Hbf/HbA1C without changing any column/cartridge |
| Page 66, Para 18 | It should have an 80GB hard disk and a remote data access feature when connected to LAN or Intranet | It should have an 80GB hard disk internal/external with a remote data access feature when connected to LAN or Intranet. |
| Page 66, Para 19 | The company should be able to provide normal and abnormal controls for Hb A2, Hb F and Hb S and provide quality control program to help compare results with similar users worldwide | The company should be able to provide normal and abnormal controls for Hb A2, Hb F and Hb S (If the OEM does not have Hb S control, it can be provided from a third party) and provide quality control program to help compare results with similar users worldwide |

Note:

- i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.
- ii. All other contents of the tender enquiry including terms & conditions remain unaltered.