

Amendment no. 3

Sub: Technical Specification Amendment for the tender of Mother & Child Block -AIIMS , New Delhi-110029.

Ref.: GeM Bid No GEM/2022/B/2700653 Dated: 03-11-2022 for item 'Embedding Station (Qty: 01 no.)'

The following changes are being incorporated in the above referred GeM Bid Number only.

Sr. No.	Buyer uploaded ATC document Page & Para	EXISTING TENDER SPECIFICATION	READ AS
1	Pg 1 Para 7	Heating unit should have the temperature adjustable between 45°C to 75°C with an incremental temperature of 1°C for paraffin reservoir, cassette bath, & forceps receptacle.	Heating unit should have the temperature adjustable between 50°C to 70°C with an incremental temperature of 1°C for paraffin reservoir, cassette bath, & forceps receptacle and temp auto cut off at 75°C.
2	Pg 1 Para 8	Should have a mould warmer with temperature adjustable between 35 and 75°C.	Should have a mould warmer with temperature adjustable between 50°C to 70°C and temp auto cut off at 75°C.
3	Pg 1 Para 10	The cooling module (cryoconsole) should be ceramic coated; should be temperature adjustable between room temperature and -6° Celsius or lower.	The cooling module (cryoconsole) should be completely insulated ; should be temperature adjustable between -3°C to -12°C
4	Pg 1 Para 12	Inbuilt USB, LAN connectivity MUST be available. Wi-Fi connectivity is desirable.	Deleted
5	Pg 1 Para 15	The user department is going to install LIS in near future. Integration of the equipment to the LIS system as per the EHR standard as per the Govt of India would be the responsibility of the supplier and the vendor must provide us written undertaking of support in this regard to the user department.	Deleted

Important Note:

1. Prospective Bidders are advised to check the GeM Portal regularly prior to the closing date and time of online submission of bid.
2. The Bid Security/EMD shall be valid for 110 days from Techno – Commercial Tender opening date. The techno-commercial tender opening date shall be considered as mentioned in bid document read with it's amendment(s), if any.
3. In reference to the extension of bid opening date as mention on GeM portal, Participating bidders, who are submitting their Bid security/EMD in the form of Bank Guarantee (BG) or FDR are instructed to extend the validity of their BG/FDR accordingly, i.e. required validity of EMD is to be kept in line with the extended bid opening date.
4. Traders/resellers/distributors/authorized agents will not be considered for availing benefits under PP Policy 2012 for MSEs as per MSE guidelines issued by MoMSME
5. Bidders shall ensure that their tender(s), complete in all respects, are submitted online through **GeM portal only**.