

Response To Pre-Bid Queries (Pre-Bid date: 24.02.2022)				
Bid Number: GEM/2022/B/1942087 Dated: 10-02-2022				
Item Name: Monitor for Regional Oximetry				
Sr. No.	Buyer Specification Document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	NA	NA	<p>As per our discussion in Pre- Bid meeting for Monitor for Regional Oximetry held on 24th February 2022, I request you to kindly add this product in exemption from make in India because government has already added this product in exemption as per latest GTE under Ruke 161 (iv) of General financial Rules ( GFRs) 2017 ( Serial No,- 32 – Near Infrared Spectrometer - NIRS ) .</p> <p>Near-infrared spectroscopy (NIRS), which is a noninvasive tool, gives a venous weighted estimate of the regional cerebral oxygen saturation (rSO2). Near infrared spectrometer (NIRS) and regional oximetry both are same and synonyms of each other.</p> <p>Near infrared spectrometer OR Regional oximetry are used to determine the oxygenation level at the site of left brain and right brain during the surgery.</p> <p>So we hereby request you to consider Near infrared spectrometer (NIRS) and regional oximetry as Same and do the needful.</p>	Not considered