

Response To Pre-Bid Queries (Pre-Bid date: 22.02.2022)				
Bid Number: GEM/2022/B/1962669 Dated: 17-02-2022				
Item Name: LOW END – MID END VENTILATOR 1962669				
Sr. No.	ATC Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Pg 1 Para 3	Built-in TFT/Led colour touch screen, graphical display of at least 15" for monitoring of real time waveform, loops & monitors values.	No change considered	No change considered
			Built-in TFT/Led colour touch screen, graphical display of at least 12- 15" for monitoring of real time waveform, loops & monitors values. Remarks: Screen size 12"- 15" enable higher participation	
			Make it 12 inch or more Justification: Please ammend this point for wider participation	
2	Pg 2 Para 11 c)	Display user defined trends (tabular and graphical) for 48-72 hrs.	Request you to please amend trend from 72 hrs to 24 hr Justification: For wider participation	Amended :- Display user defined trends (tabular and graphical) for 24 hours
3	Pg 2 Para 16	Should have inbuilt ultrasonic nebulizer for nebulizing drugs without interrupting mechanical ventilation & should have EtCO2 with capnography waveform display.	Inbuilt ultrasonic to inbuilt pneumatic nebulizer. Justification: For wider participation	Amended:- Should have inbuilt pneumatic nebulizer capable of delivering <or 3 micron particle for nebulizing drugs without interrupting mechanical ventilation & should have EtCO2 with capnography waveform display.
			Should have inbuilt ultrasonic nebulizer based on vibrating mesh technology and capable for delivering <or = 3 micron. Nebulizing drugs without interrupting mechanical ventilation & should have EtCO2 with capnography waveform display. Remarks: Vibrating mesh nebulizers offers smaller and uniform particle size lead to maximum deposition of drugs in lungs	
			Make it inbuilt/integrated ultrasonic nebulizer Justification: Please ammend this point for wider participation	

Response To Pre-Bid Queries (Pre-Bid date: 22.02.2022)				
Bid Number: GEM/2022/B/1962669 Dated: 17-02-2022				
Item Name: LOW END – MID END VENTILATOR 1962669				
Sr. No.	ATC Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
4	Pg 2 Para 17	Should have servo controlled USFDA approved humidifier.	For wider participation & healthier competition We request to amend US FDA/ EU CE	Amended as:- Ventilator and compressor should be European CE with four digit notified body number or US FDA or BIS approved for the quoted model and certificate to be submitted. OR Should meet IEC 60601-1, IEC 60601-2 & ISO 80601-2-12 standards and should submit valid test report from NABL accredited labs or from the labs in their country of origin (incase of forign maufacture) for the quoted model.
5	Pg 3 Para 26	Should be capable of upgrading for charting solutions.	Should be deleted Remarks: Enable higher participation Should be removed Justification: For wider wider participation For wider participation & healthier competition We request to Delete this point.	Amended as:- DELETED

Response To Pre-Bid Queries (Pre-Bid date: 22.02.2022)				
Bid Number: GEM/2022/B/1962669 Dated: 17-02-2022				
Item Name: LOW END – MID END VENTILATOR 1962669				
Sr. No.	ATC Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
6	Pg 26 Para 4	Preference to Make In India products (For bids less than 200 Crore):Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document 50%. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as guideline issued by Department of Pharmaceuticals vide Ref. 31026/36/2016-MD dtd 16.02.2021. Only Class-I and Class-II Local suppliers as per guideline issued by Department of Pharmaceuticals vide Ref. 31026/36/2016-MD dtd 16.02.2021 will be eligible to bid. Non - Local suppliers as per MII guideline issued by Department of Pharmaceuticals vide Ref. 31026/36/2016-MD dtd 16.02.2021 are not eligible to participate. In case Buyer has selected Purchase	Make it 25% for Class II Local supplier Justification: For wider wider participation	No change considered Clarified as:- Class I & Class II vendors are eligible to quote