

Response To Pre-Bid Queries (Pre-Bid date: 24.02.2022)					
Bid Number: GEM/2022/B/1933858 Dated: 14-02-2022					
Item Name: Plasmapheresis Machine (Portable) 1933858					
Sr. No.	Buyer Specification Document Page & Para	TENDER SPECIFICATION	NAME OF THE FIRM	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Pg 1 Para 14	The system should provide blood prime recommendations for the safety of pediatric and anemic patients	NA	As per discussion in prebid	Amended as 'The system should have provision for blood prime for the safety of pediatric and anemic patients'
2	Pg 1 Para 15	The system should have provision to initiate boluses and record the volumes.	NA	As per discussion in prebid	Amended as "The system should have provision to provide boluses or infusion through return line"
3	Pg 1 Para 16	The system should provide a continuous graphical representation of details of ongoing fluid balance	NA	As per discussion in prebid	Amended as 'The system should provide a continuous graphical or numerical representation of details of ongoing fluid balance'
4	Pg 1 Para 18	The system should have an in-built tube sealer.	NA	As per discussion in prebid	Amended as "The system should have either an inbuilt tube sealer or the option to add on (optional)"
5	Pg 1 Para 23	The system should be able to regulate pump speeds based on type of replacement solution.	NA	As per discussion in prebid	Amended as 'The system should be able to regulate pump speeds to alter the infusion rate of solution'
6	Pg 2 Para 25	The system should be able to operate in ambient temperature of 20-30°C and relative humidity 0-70%.	NA	As per discussion in prebid	Amended as 'The system should be able to operate in ambient temperature of 20-25°C and relative humidity 0-70%'
7	Pg 2 Para 32 (vi)	Machine downtime 48 hours, with a penalty of Rs. 5000/- every day after downtime	NA	As per discussion in prebid	No change
8	Pg 2 Para 35	Rates of all consumables should be provided and fixed for a period of 10 years.	NA	As per discussion in prebid	Amended to 'List of all essential consumables for use with machine to be provided as annexure (attached), and rates of all these consumables should be provided and fixed for a period of 10 years.'