

Response To Pre-Bid Queries (Pre-Bid date: 24.02.2022)				
Bid Number: GEM/2022/B/1955211 Dated: 15-02-2022				
Item Name: Measure through Motion & Low Perfusion Pulse Oximeter with NIBP 1955211				
Sr. No.	Buyer Specification Document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	NA	Suggested by Bidder	Data transfer software must be provided with the device.	Added Para Data transfer software must be provided with the device.
2	NA	Suggested by Bidder	Data transfer cable must be provided with the device.	Added Para Data transfer cable must be provided with the device.
3	NA	Suggested by Bidder	It should have stand.	Added Para Each machine should have stand.
4	NA	Suggested by Bidder	It should have 30 reusable and 30 disposable NIBP cuff.	Added Para Each machine should be supplied with 30 reusable and 30 disposable NIBP cuff.
5	NA	Suggested by Bidder	It should have 5 hose pipe for NIBP	Added Para Each Machine should be supplied with 5 hose pipe for NIBP