

Response To Pre-Bid Queries (Pre-Bid date: 24.02.2022)

Bid Number: GEM/2022/B/1952018 dated 14.02.2022

Item Name: CRRT Machine 1952018

Sr. No.	ATC Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Pg 1 Para 7	There should be provision of separate pump for pre-blood infusion	Kindly amend it as deleted. Justification: Brand-specific	Amended to 'There should be provision of separate fifth pump for pre-blood infusion'
2	Pg 1 Para 11	The system should have provision of separate weighing scale for pre blood pump	Kindly amend it as deleted. Justification: Brand-specific	No change
3	Pg 1 Para 16	The system should have provision for regional citrate anticoagulation for all modalities of therapy	Kindly amend it as deleted. Justification: Brand-specific	No change
4	Pg 2 Para 21	The system should have provision of changing syringe size	The system should have the provision of 50 ml syringe size. Justification: Treatment usually runs for 24 to 72hours, where we require a large volume of systemic Heparin syringes like 50ml. There is no use of syringes change option	Amended to 'The system should have provision of changing syringe size (optional) '
5	Pg 2 Para 23	The system should have an inbuilt dosage calculator	Kindly amend it as optional Justification: The dose calculator does not calculate downtime during treatment, which leads to a difference between the prescribed dose and the delivered dose. Therapy outcomes always depend on the prescribed dose only.	Amended to 'The system should have an inbuilt dosage calculator (optional) '
6	Pg 2 Para 35	The system should have provision for recording details of therapy for up to 80 -90 hours, and at least 500 events	The system should have provision for recording details of therapy for upto 72 hours, and at least 500 events. Justification: A single CRRT treatment usually runs for 24 to 72 hours, Max.	Amended to 'The system should have provision for recording details of therapy for at least 48 hours , and at least 100 events'

Response To Pre-Bid Queries (Pre-Bid date: 24.02.2022)

Bid Number: GEM/2022/B/1952018 dated 14.02.2022

Item Name: CRRT Machine 1952018

Sr. No.	ATC Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
7	Pg 3 Para 43	The system should have battery back up of at least 60 minutes	The system should have battery backup of at least 15 minutes Justification: Battery backup in equipment helps the user to return the blood back or to terminate the treatment in case of power failure. We have 15min battery backup in multifiltrate by which the user can return the blood	Amended to 'The system should have battery back up of at least 15 minutes '

COMMERCIAL QUERY

Sr. No.	ATC Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Pg 26 Para 4	Preference to Make In India products (For bids less than 200 Crore):Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document 50%. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as guideline issued by Department of Pharmaceuticals vide Ref. 31026/36/2016-MD dtd 16.02.2021. Only Class-I and Class-II Local suppliers as per guideline issued by Department of Pharmaceuticals vide Ref. 31026/36/2016-MD dtd 16.02.2021 will be eligible to bid. Non - Local suppliers as per MII guideline issued by Department of Pharmaceuticals vide Ref. 31026/36/2016-MD dtd 16.02.2021 are not eligible to participate. In case Buyer has selected Purchase preference to Micro and Small Enterprises clause in the bid, the same will get precedence over this clause.	Please refer to the notice No. 35026/176/2021-MD, dated 1st February 2022, issued by Government of India, Ministry of Chemicals and Fertilizers, Department of Pharmaceuticals. As per the same, relaxation has been given under rule 161 (IV) of GFR 2017 to 391 Medical Devices on the ground of non-availability of Class-I or Class-II suppliers. The CRRT machine is mentioned at S. No 162 in the Govt order list. Hence we request you to please change the MII preference as mentioned in the tender document to ensure multiple bidders can participate in the said tender. Kindly amend it as Non-availability of Local suppliers. System optional restrictions on GeM portal	No change considered