

Response To Pre-Bid Queries (Pre-Bid date: 03.10.2022)				
Bid Number: GEM/2022/B/2577791 Dated: 26-09-2022				
Defibrillator with disposable pads & pacing facility with feedback				
Sr. No.	Buyer uploaded ATC document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Pg 1 Para 9	It should have ability to measure chest compression rate and depth in real time with visual feedback on screen with rate and depth indicator.	<p>Kindly delete this point. Justification: Kindly note that the asked specification is available with only few companies which is limiting the competition. Hence, we request you to kindly delete this point for wider participation.</p> <p>To be Delete {for wider participation)</p>	Deleted
2	Pg 1 Para 14 d.	12 lead	<p>5/6 lead Justification: Kindly amend this point as the asked specification is available with one or two companies which is restricting and limiting the competition; for wider and better participation we request you to amend this point as requested</p>	Amended as: 5/6 lead
3	Pg 1 Para 15 e)	Reusable CPR feedback sensor/or similar product reused at least on 90 patients – 2 nos	<p>Kindly delete this point. Justification: Asked specification is limiting the competition so for wider and healthier participation kindly Delete this point.</p>	Deleted