

Response To Pre-Bid Queries (Pre-Bid date: 03.10.2022)				
Bid Number: GEM/2022/B/2578325 Dated: 26-09-2022				
Video Laryngoscope with Adult Blades (Reusable)				
Sr. No.	Buyer uploaded ATC document Page & Para	TENDER SPECIFICATION	REPRESENTATION RECEIVED FROM THE FIRM	COMMITTEE RECOMMENDATION
1	Pg 1 Para 1	Macintosh blade with closed European Metal finish size 3 and 4 with integrated camera chip and LED is required	Please specify the total no. of blades required & Material of the blades A. Material should be added as medical grade St. Steel SUS 316L and total no of blades required can be selectable from the following available sizes MAC - 1,2,3,4,5 MILLER - 00,0,1 MAC D Hyperangulated blades for difficult intubation - 1,2,3,4,5	No Change
2	Pg 1 Para 2	One special adult blade for difficult intubation with angle of view approx 80 degree should be provided.		
3	Pg 1 Para 3	All blades should be reusable and metallic.		
4	Pg 1 Para 4	Screen at least 7inch or more in size for display with feature control buttons on the screen with HDMI output for connecting to a big screen.		
5	Pg 1 Para 6	Monitor should have a facility to connect flexible scope and video-laryngoscope blade at same time without any special coupler.	Monitor should have a facility to connect flexible scope and video laryngoscope blade at the same time. The above should be changed to separate monitor for Video Laryngoscope and flexible scopes should be provided.	No Change
6	Pg 1 Para 17	Blades and connection cable should be fully immiscible in disinfecting solution	Blades and connecting cable should be fully immiscible in disinfection solution. It should be changed to as per details given below Only Blades should be fully liquid immersible for complete disinfection & the blades should be IPX8 compliant.	No Change
7	Pg 1 Para 20	It should be USFDA / European CE with four digit notified body / BIS certified.	It should be USFDA / European CE with 04 digit notified body / BIS certified It should be changed to USFDA registered / EC Declaration of Conformity as this is medical device class 1A & no notified body involvement is required for medical device falling under medical devices class 1A (Only self Declaration is required)	No Change
8	Pg 1 Para 22	Price of extra blade should be quoted separately.	It is possible only if GEM portal allows the option of prices of the extra blades to be quoted separately.	Deleted